

ABX Minocal



LOT CX 081
Rev 1

CAL

(Exp.) 2011-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse										TOLERANCES TOLERANCE
		MICROS					PENTRA					
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80		
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	9.2	9.2	9.0	9.1	9.3		9.3	9.3	9.2	± 0.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.65	4.63	4.62	4.65	4.65		4.66	4.65	4.63	± 0.06	
HB HGB	g/dl	13.6	13.8	13.7	13.6	13.7		13.6	13.7	13.7	± 0.2	
	g/l	136	138	137	136	137		136	137	137	± 2	
	mmol/l	8.42	8.57	8.51	8.45	8.51		8.45	8.51	8.51	± 0.12	
HT HCT	%	40.6	39.8	40.1	39.7	40.2		38.4	38.1	38.1	± 1.0	
	l/l	0.406	0.398	0.401	0.397	0.402		0.384	0.381	0.381	± 0.010	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	239	239	240	242	244		242	245	245	± 10	
VPM MPV	µm ³ ; fl	8.3	8.2	N/A	8.5	8.3		N/A	N/A	N/A	± 0.5	

Ref: FORM-0387 Rev.19 FRONT / RECTO 9930053-B

ABX Minocal



LOT CX 081
Rev 1

CAL

(Exp.) 2011-09-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l						9.1	9.3	9.3	9.2	± 0.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l						4.62	4.66	4.65	4.63	± 0.06
HB HGB	g/dl						13.6	13.6	13.7	13.7	± 0.2
	g/l						136	136	137	137	± 2
	mmol/l						8.45	8.45	8.51	8.51	± 0.12
HT HCT	%						37.4	38.4	38.1	38.1	± 1.0
	l/l						0.374	0.384	0.381	0.381	± 0.010
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l						245	242	245	245	± 10
VPM MPV	µm ³ ; fl						N/A	N/A	N/A	N/A	± 0.5

Ref: FORM-0387 Rev.19 BACK / VERSO 9930053-B