

ABX Minocal



LOT

CX 056

CAL



(Exp.) 2016-06-05

(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|-----------------|----------------|--------------|--------------|-----------------------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP CRP 200 | 8P | 18P | 60 8P 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.2 | 9.0 | 9.1 | 9.1 | 9.1 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.65 | 4.59 | 4.66 | 4.65 | 4.63 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.7 | 13.7 | 13.7 | 13.8 | 13.8 | | | | | ± 0.2 |
| | g/l | 137 | 137 | 137 | 138 | 138 | | | | | ± 2 |
| | mmol/l | 8.51 | 8.51 | 8.51 | 8.57 | 8.57 | | | | | ± 0.12 |
| HT HCT | % | 39.3 | 38.8 | 39.4 | 39.8 | 39.1 | | | | | ± 1.0 |
| | l/l | 0.393 | 0.388 | 0.394 | 0.398 | 0.391 | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 249 | 245 | 246 | 248 | 244 | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | 9.0 | 8.6 | N/A | 9.0 | 8.3 | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.42 FRONT / RECTO 1219930090 -A

ABX Minocal



LOT

CX 056

CAL



(Exp.) 2016-06-05
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|--------|--------------|--------------|----------------------|--------------|----------------------------------|---------------------|--------------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | 120 - 120 RETIC DX120 - DF120 | 60 60 C+ ES60 | 80 XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.0 | | 9.4 | 9.2 | 9.3 | 9.2 | 9.3 | 9.3 | 9.2 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.60 | | 4.54 | 4.60 | 4.64 | 4.60 | 4.64 | 4.60 | 4.60 | ± 0.06 |
| HB HGB | g/dl | 13.5 | | 13.5 | 13.6 | 13.6 | 13.6 | 13.6 | 13.6 | 13.7 | ± 0.2 |
| | g/l | 135 | | 135 | 136 | 136 | 136 | 136 | 136 | 137 | ± 2 |
| | mmol/l | 8.38 | | 8.38 | 8.45 | 8.45 | 8.45 | 8.45 | 8.45 | 8.51 | ± 0.12 |
| HT HCT | % | 38.6 | | 36.5 | 37.0 | 37.9 | 36.8 | 37.9 | 37.1 | 37.5 | ± 1.0 |
| | l/l | 0.386 | | 0.365 | 0.370 | 0.379 | 0.368 | 0.379 | 0.371 | 0.375 | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 246 | | 230 | 235 | 248 | 252 | 248 | 248 | 246 | ± 10 |
| VPM MPV | µm ³ ; fl | 8.8 | | N/A | N/A | N/A | N/A | N/A | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.42 BACK / VERSO 1219930090 -A