

LOT CX 124
Rev 1

CAL

 (Exp.) 2015-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.5	9.2	9.2	9.3	9.4					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.53	4.50	4.55	4.55	4.52					± 0.06
HB HGB	g/dl	13.9	13.8	13.7	13.8	13.8					± 0.2
	g/l	139	138	137	138	138					± 2
	mmol/l	8.63	8.57	8.51	8.57	8.57					± 0.12
HT HCT	%	39.0	38.8	39.4	39.1	39.1					± 1.0
	l/l	0.390	0.388	0.394	0.391	0.391					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	240	234	236	240	238					± 10
VPM MPV	µm ³ ; fl	8.8	8.7	N/A	9.1	8.0					± 0.5

Ref: TEMP-0387 Rev.38 FRONT / RECTO 9930090 -A

LOT CX 124
Rev 1

CAL

 (Exp.) 2015-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP		MS CRP	XLR	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.4		9.2	9.2	9.5	9.2	9.5	9.2	9.2	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.49		4.47	4.54	4.54	4.56	4.54	4.55	4.54	± 0.06
HB HGB	g/dl	13.5		13.5	13.7	13.7	13.8	13.7	13.7	13.7	± 0.2
	g/l	135		135	137	137	138	137	137	137	± 2
	mmol/l	8.38		8.38	8.51	8.51	8.57	8.51	8.51	8.51	± 0.12
HT HCT	%	38.2		37.0	37.3	37.5	37.0	37.5	37.2	37.3	± 1.0
	l/l	0.382		0.370	0.373	0.375	0.370	0.375	0.372	0.373	± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	240		236	236	235	236	235	237	236	± 10
VPM MPV	µm ³ ; fl	8.9		N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.38 BACK / VERSO 9930090 -A