

**LOT** CX 084  
Rev 1

**CAL**

 (Exp.) 2014-09-05  
(YYYY - MM - DD)

| PARAMETRES<br>PARAMETERS | UNITES<br>UNITS                                       | ABX Lyse        |                |        |        |                             |        |        |        | TOLERANCES<br>TOLERANCE |         |
|--------------------------|---|-----------------|----------------|--------|--------|-----------------------------|--------|--------|--------|-------------------------|---------|
|                          |   | MICROS          | MICROS         | MICROS | MICROS | MICROS                      | PENTRA | PENTRA | PENTRA |                         | PENTRA  |
|                          |   | ES60<br>Care ST | CRP<br>CRP 200 | 8P     | 18P    | 60 8P<br>60 18P<br>ADVIA 60 |        |        |        |                         |         |
| GB WBC                   | 10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l  | 9.3             | 9.2            | 9.1    | 9.1    | 9.2                         |        |        |        |                         | ± 0.2   |
| GR RBC                   | 10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l | 4.45            | 4.42           | 4.43   | 4.46   | 4.46                        |        |        |        |                         | ± 0.06  |
| HB HGB                   | g/dl  | 13.7            | 13.7           | 13.6   | 13.7   | 13.7                        |        |        |        |                         | ± 0.2   |
|                          | g/l   | 137             | 137            | 136    | 137    | 137                         |        |        |        |                         | ± 2     |
|                          | mmol/l  | 8.51            | 8.51           | 8.45   | 8.51   | 8.51                        |        |        |        |                         | ± 0.12  |
| HT HCT                   | %   | 38.7            | 38.7           | 38.5   | 38.8   | 39.0                        |        |        |        |                         | ± 1.0   |
|                          | l/l   | 0.387           | 0.387          | 0.385  | 0.388  | 0.390                       |        |        |        |                         | ± 0.010 |
| PLAQ. PLTS               | 10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l  | 259             | 250            | 255    | 254    | 251                         |        |        |        |                         | ± 10    |
| VPM MPV                  | µm <sup>3</sup> ; fl                                  | 8.3             | 8.0            | N/A    | 8.5    | 7.6                         |        |        |        |                         | ± 0.5   |

Ref: TEMP-0387 Rev.37 FRONT / RECTO 9930090 -A

**LOT** CX 084  
Rev 1

**CAL**

 (Exp.) 2014-09-05  
(YYYY - MM - DD)

| PARAMETRES<br>PARAMETERS | UNITES<br>UNITS                                       | ABX Lysebio |        |        |        |                      |        |                                  |                     |            | TOLERANCES<br>TOLERANCE |
|--------------------------|---|-------------|--------|--------|--------|----------------------|--------|----------------------------------|---------------------|------------|-------------------------|
|                          |   | MICROS      | MICROS | PENTRA | PENTRA | PENTRA               | PENTRA | PENTRA                           | PENTRA              | PENTRA     |                         |
|                          |   | EMI CRP     |        | MS CRP | XLR    | DX NEXUS<br>DF NEXUS | MS60   | 120 - 120 RETIC<br>DX120 - DF120 | 60<br>60 C+<br>ES60 | 80<br>XL80 |                         |
| GB WBC                   | 10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l  | 9.2         |        | 9.0    | 9.0    | 9.3                  | 9.0    | 9.3                              | 9.1                 | 9.0        | ± 0.2                   |
| GR RBC                   | 10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l | 4.41        |        | 4.35   | 4.43   | 4.46                 | 4.43   | 4.46                             | 4.44                | 4.43       | ± 0.06                  |
| HB HGB                   | g/dl  | 13.5        |        | 13.5   | 13.6   | 13.6                 | 13.6   | 13.6                             | 13.6                | 13.6       | ± 0.2                   |
|                          | g/l   | 135         |        | 135    | 136    | 136                  | 136    | 136                              | 136                 | 136        | ± 2                     |
|                          | mmol/l  | 8.38        |        | 8.38   | 8.45   | 8.45                 | 8.45   | 8.45                             | 8.45                | 8.45       | ± 0.12                  |
| HT HCT                   | %   | 37.9        |        | 36.5   | 37.1   | 37.4                 | 36.3   | 37.4                             | 37.1                | 37.1       | ± 1.0                   |
|                          | l/l   | 0.379       |        | 0.365  | 0.371  | 0.374                | 0.363  | 0.374                            | 0.371               | 0.371      | ± 0.010                 |
| PLAQ. PLTS               | 10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l  | 251         |        | 258    | 252    | 258                  | 250    | 258                              | 253                 | 252        | ± 10                    |
| VPM MPV                  | µm <sup>3</sup> ; fl                                  | 8.2         |        | N/A    | N/A    | N/A                  | N/A    | N/A                              | N/A                 | N/A        | ± 0.5                   |

Ref: TEMP-0387 Rev.37 BACK / VERSO 9930090 -A