

ABX Minocal



LOT

CX 423

CAL



(Exp.) 2018-07-05

(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|-----------------|----------------|--------------|--------------|-----------------------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP CRP 200 | 8P | 18P | 60 8P 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.1 | 9.1 | 8.9 | 9.0 | 9.0 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.42 | 4.42 | 4.43 | 4.45 | 4.43 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.6 | 13.7 | 13.7 | 13.6 | 13.7 | | | | | ± 0.2 |
| | g/l | 136 | 137 | 137 | 136 | 137 | | | | | ± 2 |
| | mmol/l | 8.45 | 8.51 | 8.51 | 8.45 | 8.51 | | | | | ± 0.12 |
| HT HCT | % | 38.5 | 38.7 | 39.0 | 39.2 | 39.0 | | | | | ± 1.0 |
| | l/l | 0.385 | 0.387 | 0.390 | 0.392 | 0.390 | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 248 | 244 | 235 | 250 | 250 | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | 8.6 | 8.6 | N/A | 9.1 | 7.9 | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.45 FRONT / RECTO 1219930090 -A

ABX Minocal



LOT

CX 423

CAL



(Exp.) 2018-07-05
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|--------|--------------|--------------|----------------------|--------------|----------------------------------|---------------------|--------------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | 120 - 120 RETIC DX120 - DF120 | 60 60 C+ ES60 | 80 XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.1 | | 9.0 | 9.0 | 9.2 | 8.9 | 9.2 | 9.1 | 9.0 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.42 | | 4.44 | 4.41 | 4.46 | 4.45 | 4.46 | 4.43 | 4.41 | ± 0.06 |
| HB HGB | g/dl | 13.6 | | 13.7 | 13.6 | 13.7 | 13.7 | 13.7 | 13.7 | 13.7 | ± 0.2 |
| | g/l | 136 | | 137 | 136 | 137 | 137 | 137 | 137 | 137 | ± 2 |
| | mmol/l | 8.45 | | 8.51 | 8.45 | 8.51 | 8.51 | 8.51 | 8.51 | 8.51 | ± 0.12 |
| HT HCT | % | 38.5 | | 37.1 | 37.0 | 37.9 | 36.9 | 37.7 | 37.2 | 37.3 | ± 1.0 |
| | l/l | 0.385 | | 0.371 | 0.370 | 0.379 | 0.369 | 0.377 | 0.372 | 0.373 | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 242 | | 242 | 243 | 245 | 253 | 245 | 246 | 243 | ± 10 |
| VPM MPV | µm ³ ; fl | 8.9 | | N/A | N/A | N/A | N/A | N/A | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.45 BACK / VERSO 1219930090 -A

ABX Minocal



LOT

CX 423

CAL



(Exp.) 2018-07-05
(YYYY-MM-DD)

Rev **1**

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|----------------|--|--|--|--|--|--|--|-------------------------|
| | | YUMIZEN | | | | | | | | |
| | | H1500 H2500 | | | | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 8.8 | | | | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.48 | | | | | | | | ± 0.06 |
| HB HGB | g/dl | 13.6 | | | | | | | | ± 0.2 |
| | g/l | 136 | | | | | | | | ± 2 |
| | mmol/l | 8.45 | | | | | | | | ± 0.12 |
| HT HCT | % | N/A | | | | | | | | N/A |
| | l/l | N/A | | | | | | | | N/A |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 263 | | | | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | N/A | | | | | | | | N/A |
| VGM MCV | µm ³ ; fl | 84.5 | | | | | | | | ± 4.0 |

1300047683
Ref: TEMP-0387 Rev.45