

ABX Minocal



LOT

CX 462

CAL



(Exp.) 2021-10-05

Rev 1

(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|-----------------|--------------|--------|--------|-----------------------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP 200 | | | 60 8P 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.4 | 9.0 | | | 9.2 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.55 | 4.60 | | | 4.52 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.7 | 13.8 | | | 13.7 | | | | | ± 0.2 |
| | g/l | 137 | 138 | | | 137 | | | | | ± 2 |
| | mmol/l | 8.51 | 8.57 | | | 8.51 | | | | | ± 0.12 |
| HT HCT | % | 39.1 | 39.3 | | | 38.9 | | | | | ± 1.0 |
| | l/l | 0.391 | 0.393 | | | 0.389 | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 255 | 250 | | | 245 | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | 8.7 | 8.7 | | | 8.1 | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.47 FRONT / RECTO 1300090308

ABX Minocal



LOT

CX 462

CAL



(Exp.) 2021-10-05
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|--------------|--------------|--------------|----------------------|--------------|---------------|---------------------|--------------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | LC-767G | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | DX120 - DF120 | 60 60 C+ ES60 | XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.3 | 9.2 | 9.4 | 9.2 | 9.7 | 9.2 | 9.5 | 9.2 | 9.2 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.53 | 4.50 | 4.45 | 4.53 | 4.56 | 4.52 | 4.56 | 4.55 | 4.53 | ± 0.06 |
| HB HGB | g/dl | 13.6 | 13.4 | 13.5 | 13.6 | 13.6 | 13.5 | 13.6 | 13.7 | 13.6 | ± 0.2 |
| | g/l | 136 | 134 | 135 | 136 | 136 | 135 | 136 | 137 | 136 | ± 2 |
| | mmol/l | 8.45 | 8.32 | 8.38 | 8.45 | 8.45 | 8.38 | 8.45 | 8.51 | 8.45 | ± 0.12 |
| HT HCT | % | 38.7 | 38.5 | 36.9 | 37.6 | 37.4 | 36.8 | 37.6 | 37.8 | 37.6 | ± 1.0 |
| | l/l | 0.387 | 0.385 | 0.369 | 0.376 | 0.374 | 0.368 | 0.376 | 0.378 | 0.376 | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 255 | 248 | 244 | 242 | 251 | 247 | 251 | 248 | 242 | ± 10 |
| VPM MPV | µm ³ ; fl | 8.9 | 8.8 | N/A | N/A | N/A | N/A | N/A | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.47 BACK / VERSO 1300090308

ABX Minocal



LOT

CX 462

CAL



(Exp.) 2021-10-05

(YYYY-MM-DD)

Rev **1**

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|----------------|--|--|--|--|--|--|--|-------------------------|
| | | YUMIZEN | | | | | | | | |
| | | H1500 H2500 | | | | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 8.80 | | | | | | | | ± 0.20 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.57 | | | | | | | | ± 0.06 |
| HB HGB | g/dl | 13.6 | | | | | | | | ± 0.2 |
| | g/l | 136 | | | | | | | | ± 2 |
| | mmol/l | 8.45 | | | | | | | | ± 0.12 |
| HT HCT | % | N/A | | | | | | | | N/A |
| | l/l | N/A | | | | | | | | N/A |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 251 | | | | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | N/A | | | | | | | | N/A |
| VGM MCV | µm ³ ; fl | 83.5 | | | | | | | | ± 4.0 |

1300060884

Ref: TEMP-0387 Rev.47