

ABX Minocal



LOT

CX 468

CAL



(Exp.) 2022-04-05

Rev 1

(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|-----------------|--------------|--------|--------|-----------------------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP 200 | | | 60 8P 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.3 | 9.2 | | | 9.0 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.54 | 4.62 | | | 4.55 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.7 | 13.9 | | | 13.8 | | | | | ± 0.2 |
| | g/l | 137 | 139 | | | 138 | | | | | ± 2 |
| | mmol/l | 8.51 | 8.63 | | | 8.57 | | | | | ± 0.12 |
| HT HCT | % | 38.8 | 40.0 | | | 39.4 | | | | | ± 1.0 |
| | l/l | 0.388 | 0.400 | | | 0.394 | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 252 | 255 | | | 251 | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | 8.9 | 8.5 | | | 7.9 | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.47 FRONT / RECTO 1300090308

ABX Minocal



LOT

CX 468

CAL



(Exp.) 2022-04-05
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|-------------|---------|--------|--------|----------------------|--------|---------------|---------------------|--------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | LC-767G | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | DX120 - DF120 | 60 60 C+ ES60 | XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.1 | 9.2 | 9.1 | 9.2 | 9.5 | 9.0 | 9.2 | 9.2 | 9.2 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.52 | 4.55 | 4.50 | 4.52 | 4.60 | 4.57 | 4.60 | 4.55 | 4.52 | ± 0.06 |
| HB HGB | g/dl | 13.6 | 13.6 | 13.6 | 13.9 | 13.8 | 13.7 | 13.8 | 13.8 | 13.9 | ± 0.2 |
| | g/l | 136 | 136 | 136 | 139 | 138 | 137 | 138 | 138 | 139 | ± 2 |
| | mmol/l | 8.45 | 8.45 | 8.45 | 8.63 | 8.57 | 8.51 | 8.57 | 8.57 | 8.63 | ± 0.12 |
| HT HCT | % | 38.4 | 38.7 | 37.4 | 37.5 | 37.7 | 37.2 | 37.7 | 37.5 | 37.5 | ± 1.0 |
| | l/l | 0.384 | 0.387 | 0.374 | 0.375 | 0.377 | 0.372 | 0.377 | 0.375 | 0.375 | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 256 | 258 | 246 | 248 | 251 | 247 | 251 | 245 | 248 | ± 10 |
| VPM MPV | µm ³ ; fl | 8.9 | 8.8 | N/A | N/A | N/A | N/A | N/A | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.47 BACK / VERSO 1300090308

ABX Minocal



LOT

CX 468

CAL



(Exp.) 2022-04-05

(YYYY-MM-DD)

Rev **1**

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|----------------|--|--|--|--|--|--|--|-------------------------|
| | | YUMIZEN | | | | | | | | |
| | | H1500 H2500 | | | | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.00 | | | | | | | | ± 0.20 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.64 | | | | | | | | ± 0.06 |
| HB HGB | g/dl | 14.0 | | | | | | | | ± 0.2 |
| | g/l | 140 | | | | | | | | ± 2 |
| | mmol/l | 8.69 | | | | | | | | ± 0.12 |
| HT HCT | % | N/A | | | | | | | | N/A |
| | l/l | N/A | | | | | | | | N/A |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 260 | | | | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | N/A | | | | | | | | N/A |
| VGM MCV | µm ³ ; fl | 84.0 | | | | | | | | ± 4.0 |

1300060884

Ref: TEMP-0387 Rev.47