

**LOT** CX 123  
Rev 1

**CAL**

(Exp.) 2014-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	9.5	9.3	9.2	9.2	9.4			9.4	9.3	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	4.51	4.50	4.50	4.48	4.50			4.53	4.51	± 0.06
HB HGB	g/dl	13.6	13.5	13.4	13.4	13.6			13.6	13.6	± 0.2
	g/l	136	135	134	134	136			136	136	± 2
	mmol/l	8.45	8.38	8.32	8.32	8.45			8.45	8.45	± 0.12
HT HCT	%	38.6	38.7	38.7	38.5	38.7			37.4	37.3	± 1.0
	l/l	0.386	0.387	0.387	0.385	0.387			0.374	0.373	± 0.010
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	252	243	245	253	246			245	249	± 10
VPM MPV	µm <sup>3</sup> ; fl	8.0	7.9	N/A	8.1	7.6			N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.33 FRONT / RECTO 9930081 -B

**LOT** CX 123  
Rev 1

**CAL**

 (Exp.) 2014-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	CRP CRP 200	8P	XLR	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	9.2			9.3	9.5	9.2	9.5	9.4	9.3	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	4.50			4.51	4.52	4.50	4.52	4.53	4.51	± 0.06
HB HGB	g/dl	13.5			13.6	13.5	13.6	13.5	13.6	13.6	± 0.2
	g/l	135			136	135	136	135	136	136	± 2
	mmol/l	8.38			8.45	8.38	8.45	8.38	8.45	8.45	± 0.12
HT HCT	%	38.3			37.3	37.5	36.2	37.5	37.4	37.3	± 1.0
	l/l	0.383			0.373	0.375	0.362	0.375	0.374	0.373	± 0.010
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	246			249	248	252	248	245	249	± 10
VPM MPV	µm <sup>3</sup> ; fl	8.1			N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.33 BACK / VERSO 9930081 -B