

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**CX 023**

**CAL**

(Exp.) **2013-03-05**  
(YYYY - MM - DD)

Rev **1**

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>9.8</b>	<b>9.5</b>	<b>9.6</b>	<b>9.5</b>	<b>9.7</b>		<b>9.8</b>	<b>9.7</b>	<b>9.7</b>	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.39</b>	<b>4.40</b>	<b>4.38</b>	<b>4.40</b>	<b>4.40</b>		<b>4.43</b>	<b>4.42</b>	<b>4.38</b>	± 0.06
HB HGB	g/dl	<b>13.7</b>	<b>13.6</b>	<b>13.6</b>	<b>13.6</b>	<b>13.5</b>		<b>13.6</b>	<b>13.5</b>	<b>13.5</b>	± 0.2
	g/l	<b>137</b>	<b>136</b>	<b>136</b>	<b>136</b>	<b>135</b>		<b>136</b>	<b>135</b>	<b>135</b>	± 2
	mmol/l	<b>8.51</b>	<b>8.45</b>	<b>8.45</b>	<b>8.45</b>	<b>8.38</b>		<b>8.45</b>	<b>8.38</b>	<b>8.38</b>	± 0.12
HT HCT	%	<b>38.6</b>	<b>38.7</b>	<b>38.8</b>	<b>38.9</b>	<b>38.8</b>		<b>37.5</b>	<b>37.0</b>	<b>37.1</b>	± 1.0
	l/l	<b>0.386</b>	<b>0.387</b>	<b>0.388</b>	<b>0.389</b>	<b>0.388</b>		<b>0.375</b>	<b>0.370</b>	<b>0.371</b>	± 0.010
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>236</b>	<b>236</b>	<b>238</b>	<b>240</b>	<b>233</b>		<b>238</b>	<b>240</b>	<b>238</b>	± 10
VPM MPV	µm <sup>3</sup> ; fl	<b>8.2</b>	<b>8.3</b>	N/A	<b>8.6</b>	<b>7.9</b>		N/A	N/A	N/A	± 0.5

ref: TEMP-0387 Rev:30 FRONT / RECTO 9930077 -A

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

**LOT**

**CX 023**

**CAL**

 **(Exp.) 2013-03-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	CRP CRP 200	8P	18P	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>9.3</b>				<b>9.8</b>	<b>9.6</b>	<b>9.8</b>	<b>9.7</b>	<b>9.7</b>	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.47</b>				<b>4.43</b>	<b>4.40</b>	<b>4.43</b>	<b>4.42</b>	<b>4.38</b>	± 0.06
HB HGB	g/dl	<b>13.2</b>				<b>13.6</b>	<b>13.5</b>	<b>13.6</b>	<b>13.5</b>	<b>13.5</b>	± 0.2
	g/l	<b>132</b>				<b>136</b>	<b>135</b>	<b>136</b>	<b>135</b>	<b>135</b>	± 2
	mmol/l	<b>8.20</b>				<b>8.45</b>	<b>8.38</b>	<b>8.45</b>	<b>8.38</b>	<b>8.38</b>	± 0.12
HT HCT	%	<b>39.6</b>				<b>37.5</b>	<b>36.1</b>	<b>37.5</b>	<b>37.0</b>	<b>37.1</b>	± 1.0
	l/l	<b>0.396</b>				<b>0.375</b>	<b>0.361</b>	<b>0.375</b>	<b>0.370</b>	<b>0.371</b>	± 0.010
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>233</b>				<b>238</b>	<b>235</b>	<b>238</b>	<b>240</b>	<b>238</b>	± 10
VPM MPV	µm <sup>3</sup> ; fl	<b>8.0</b>				N/A	N/A	N/A	N/A	N/A	± 0.5

ref: TEMP-0387 Rev.30 BACK / VERSO 9930077 -A