

# ABX Minocal



**LOT**

**CX 042**

**CAL**

(Exp.) **2012-05-05**  
(YYYY - MM - DD)

Rev **1**

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse										TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA		
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX 120 - DF 120	60 60 C+ ES60	80 XL80		
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>9.3</b>	<b>9.2</b>	<b>9.2</b>	<b>9.1</b>	<b>9.4</b>		<b>9.4</b>	<b>9.2</b>	<b>9.2</b>	± 0.2	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>4.53</b>	<b>4.53</b>	<b>4.50</b>	<b>4.50</b>	<b>4.53</b>		<b>4.53</b>	<b>4.50</b>	<b>4.48</b>	± 0.06	
HB HGB	g/dl	<b>13.5</b>	<b>13.8</b>	<b>13.5</b>	<b>13.5</b>	<b>13.6</b>		<b>13.6</b>	<b>13.6</b>	<b>13.6</b>	± 0.2	
	g/l	<b>135</b>	<b>138</b>	<b>135</b>	<b>135</b>	<b>136</b>		<b>136</b>	<b>136</b>	<b>136</b>	± 2	
	mmol/l	<b>8.36</b>	<b>8.57</b>	<b>8.38</b>	<b>8.38</b>	<b>8.45</b>		<b>8.45</b>	<b>8.45</b>	<b>8.45</b>	± 0.12	
HT HCT	%	<b>40.2</b>	<b>39.8</b>	<b>39.8</b>	<b>39.3</b>	<b>39.8</b>		<b>38.2</b>	<b>37.8</b>	<b>37.9</b>	± 1.0	
	l/l	<b>0.402</b>	<b>0.398</b>	<b>0.398</b>	<b>0.393</b>	<b>0.398</b>		<b>0.382</b>	<b>0.378</b>	<b>0.379</b>	± 0.010	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>240</b>	<b>240</b>	<b>243</b>	<b>240</b>	<b>245</b>		<b>240</b>	<b>245</b>	<b>244</b>	± 10	
VPM MPV	µm <sup>3</sup> ; fl	<b>7.7</b>	<b>7.7</b>	N/A	<b>7.8</b>	<b>7.7</b>		N/A	N/A	N/A	± 0.5	

Ref. FORM-0387 Rev.20 FRONT / RECTO 9930057-A

# ABX Minocal



**LOT**

**CX 042**

**CAL**

(Exp.) **2012-05-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX 120 - DF 120	60 60 C+ ES60		80 XL80
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l						9.2	9.4	9.2	9.2	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l						4.49	4.53	4.50	4.48	± 0.06
HB HGB	g/dl						13.6	13.6	13.6	13.6	± 0.2
	g/l						136	136	136	136	± 2
	mmol/l						8.45	8.45	8.45	8.45	± 0.12
HT HCT	%						36.8	38.2	37.8	37.9	± 1.0
	l/l						0.368	0.382	0.378	0.379	± 0.010
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l						236	240	245	244	± 10
VPM MPV	µm <sup>3</sup> ; fl						N/A	N/A	N/A	N/A	± 0.5

Ref: FORM-0387 Rev.20 BACK / VERSO 9930057-A