

# ABX Minocal



**LOT** CX 101  
Rev 1

**CAL**

(Exp.) 2011-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	9.3	9.4	9.3	9.3	9.4		9.5	9.4	9.3	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.57	4.57	4.60	4.55	4.57		4.59	4.58	4.56	± 0.06
HB HGB	g/dl	13.7	13.7	13.8	13.7	13.8		13.7	13.7	13.8	± 0.2
	g/l	137	137	138	137	138		137	137	138	± 2
	mmol/l	8.48	8.51	8.57	8.51	8.57		8.51	8.51	8.57	± 0.12
HT HCT	%	40.3	39.9	40.4	39.5	39.9		38.3	38.0	38.0	± 1.0
	l/l	0.403	0.399	0.404	0.395	0.399		0.383	0.380	0.380	± 0.010
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	245	245	248	251	250		247	248	248	± 10
VPM MPV	µm <sup>3</sup> ; fl	7.9	7.7	N/A	7.8	7.9		N/A	N/A	N/A	± 0.5

Ref: FORM-0387 Rev.20 FRONT / RECTO 9930057-A

# ABX Minocal



**LOT** CX 101  
Rev 1

**CAL**

(Exp.) 2011-11-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l						9.4	9.5	9.4	9.3	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l						4.55	4.59	4.58	4.56	± 0.06
HB HGB	g/dl						13.7	13.7	13.8	13.7	± 0.2
	g/l						137	137	138	137	± 2
	mmol/l						8.51	8.51	8.57	8.51	± 0.12
HT HCT	%						37.0	38.3	38.0	38.0	± 1.0
	l/l						0.370	0.383	0.380	0.380	± 0.010
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l						249	247	248	248	± 10
VPM MPV	µm <sup>3</sup> ; fl						N/A	N/A	N/A	N/A	± 0.5

Ref: FORM-0387 Rev.20 BACK / VERSO 9930057-A