

ABX Minocal



LOT CX 091
Rev 1

CAL

(Exp.) 2011-10-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	8.8	8.9	8.8	8.9	8.9		9.1	9.0	9.0	± 0.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.64	4.61	4.64	4.60		4.64	4.62	4.60	± 0.06
HB HGB	g/dl	13.5	13.8	13.7	13.6	13.6		13.6	13.6	13.7	± 0.2
	g/l	135	138	137	136	136		136	136	137	± 2
	mmol/l	8.36	8.57	8.51	8.45	8.45		8.45	8.45	8.51	± 0.12
HT HCT	%	40.4	40.0	40.0	40.1	40.0		38.4	37.8	38.0	± 1.0
	l/l	0.404	0.400	0.400	0.401	0.400		0.384	0.378	0.380	± 0.010
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	251	253	248	258	256		253	254	254	± 10
VPM MPV	µm ³ ; fl	7.9	7.7	N/A	7.9	7.9		N/A	N/A	N/A	± 0.5

Ref: FORM-0387 Rev.19 FRONT / RECTO 9930053-B

ABX Minocal



LOT CX 091
Rev 1

CAL

(Exp.) 2011-10-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l						8.8	9.1	9.0	9.0	± 0.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l						4.58	4.64	4.62	4.60	± 0.06
HB HGB	g/dl						13.6	13.6	13.6	13.7	± 0.2
	g/l						136	136	136	137	± 2
	mmol/l						8.45	8.45	8.45	8.51	± 0.12
HT HCT	%						37.4	38.4	37.8	38.0	± 1.0
	l/l						0.374	0.384	0.378	0.380	± 0.010
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l						248	253	254	254	± 10
VPM MPV	µm ³ ; fl						N/A	N/A	N/A	N/A	± 0.5

Ref: FORM-0387 Rev.19 BACK / VERSO 9930053-B