

LOT CX 014
Rev 1

CAL

(Exp.) 2014-02-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60		80 XL80
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.6	9.3	9.3	9.4	9.4			9.5	9.5	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.58	4.56	4.59	4.60	4.60			4.60	4.58	± 0.06
HB HGB	g/dl	13.7	13.5	13.5	13.6	13.6			13.6	13.6	± 0.2
	g/l	137	135	135	136	136			136	136	± 2
	mmol/l	8.51	8.38	8.38	8.45	8.45			8.45	8.45	± 0.12
HT HCT	%	38.7	38.8	39.0	39.3	39.1			37.5	37.6	± 1.0
	l/l	0.387	0.388	0.390	0.393	0.391			0.375	0.376	± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	250	243	242	253	247			247	250	± 10
VPM MPV	µm ³ ; fl	7.8	7.8	N/A	8.0	7.4			N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.33 FRONT / RECTO 9930081 -B

ABX Minocal



LOT CX 014
Rev 1

CAL

(Exp.) 2014-02-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	CRP CRP 200	8P	XLR	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.3			9.5	9.6	9.4	9.6	9.5	9.5	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.56			4.58	4.61	4.57	4.61	4.60	4.58	± 0.06
HB HGB	g/dl	13.4			13.7	13.6	13.6	13.6	13.6	13.7	± 0.2
	g/l	134			137	136	136	136	136	137	± 2
	mmol/l	8.32			8.51	8.45	8.45	8.45	8.45	8.51	± 0.12
HT HCT	%	38.3			37.6	37.8	36.5	37.8	37.5	37.6	± 1.0
	l/l	0.383			0.376	0.378	0.365	0.378	0.375	0.376	± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	240			250	247	250	247	247	250	± 10
VPM MPV	µm ³ ; fl	8.1			N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.33 BACK / VERSO 9930081 -B