

ABX Minocal

LOT
CX 115
CAL

(Exp.) 2015-12-05

(YYYY - MM - DD)

 Rev **1**

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.2	9.2	9.1	9.1	9.2				± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.51	4.50	4.53	4.55	4.52				± 0.06
HB HGB	g/dl	13.6	13.8	13.7	13.7	13.7				± 0.2
	g/l	136	138	137	137	137				± 2
	mmol/l	8.45	8.57	8.51	8.51	8.51				± 0.12
HT HCT	%	38.8	39.1	39.4	39.6	39.1				± 1.0
	l/l	0.388	0.391	0.394	0.396	0.391				± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	243	238	243	240	240				± 10
VPM MPV	µm ³ ; fl	8.8	8.6	N/A	9.0	8.1				± 0.5

Ref: TEMP-0387 Rev.40 FRONT / RECTO 1219930090 -A

ABX Minocal


 (Exp.) 2015-12-05
 (YYYY - MM - DD)

LOT		CX 115		CAL		(Exp.)2015-12-05 (YYYY - MM - DD)						
		Rev 1										
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE	
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA		
		EMI CRP		MS CRP	XLR	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80		
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.2		9.3	9.1	9.4	9.1	9.4	9.1	9.2	± 0.2	
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.48		4.50	4.54	4.55	4.54	4.55	4.55	4.54	± 0.06	
HB HGB	g/dl	13.5		13.6	13.7	13.7	13.8	13.7	13.7	13.7	± 0.2	
	g/l	135		136	137	137	138	137	137	137	± 2	
	mmol/l	8.38		8.45	8.51	8.51	8.57	8.51	8.51	8.51	± 0.12	
HT HCT	%	38.8		37.4	37.8	38.6	37.1	38.0	37.9	37.9	± 1.0	
	l/l	0.388		0.374	0.378	0.386	0.371	0.380	0.379	0.379	± 0.010	
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	240		236	234	247	244	247	240	240	± 10	
VPM MPV	µm ³ ; fl	8.8		N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5	

Ref: TEMP-0387 Rev.40 BACK / VERSO 1219930090 -A