

ABX Minocal

LOT

CX 443

CAL



(Exp.) 2020-03-05

Rev 1

(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|--------------|--------------|--------|--------|-----------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 | CRP 200 | | | 60 8P 60 18P | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.1 | 8.9 | | | 9.2 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.52 | 4.50 | | | 4.54 | | | | | ± 0.06 |
| HB HGB | g/dl | 14.1 | 13.6 | | | 14.1 | | | | | ± 0.2 |
| | g/l | 141 | 136 | | | 141 | | | | | ± 2 |
| | mmol/l | 8.76 | 8.46 | | | 8.76 | | | | | ± 0.12 |
| HT HCT | % | 39.6 | 39.4 | | | 39.7 | | | | | ± 1.0 |
| | l/l | 0.396 | 0.394 | | | 0.397 | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 250 | 235 | | | 245 | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | 8.3 | 8.1 | | | 8.8 | | | | | ± 0.5 |

ABX Minocal

LOT

CX 443

CAL



(Exp.) 2020-03-05

(YYYY-MM-DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | LyseVet | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|-----------------------------|--------|--|--|--|--|--|-------------------------|
| | | MICROS | | MICROS | | | | | | |
| | | Scil Vet ABC | Scil Vet ABC Plus ESV 60 | | | | | | | |
| GB WBC | 10 ⁹ /mm ³ ;10 ⁹ /l | 8.8 | 8.8 | | | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.51 | 4.51 | | | | | | | ± 0.06 |
| HB HGB | g/dl | 13.8 | 13.7 | | | | | | | ± 0.2 |
| | g/l | 138 | 137 | | | | | | | ± 2 |
| | mmol/l | 8.57 | 8.51 | | | | | | | ± 0.12 |
| HT HCT | % | 39.5 | 39.2 | | | | | | | ± 1.0 |
| | l/l | 0.395 | 0.392 | | | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 233 | 252 | | | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | N/A | 8.8 | | | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.46 FRONT / RECTO 1300060881