

ABX Minocal

LOT

CX 423

CAL



(Exp.) 2018-07-05

Rev 1

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	8.8	9.1			9.0					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.45	4.42			4.42					± 0.06
HB HGB	g/dl	13.8	13.4			13.7					± 0.2
	g/l	138	134			137					± 2
	mmol/l	8.57	8.34			8.51					± 0.12
HT HCT	%	38.9	38.7			38.7					± 1.0
	l/l	0.389	0.387			0.387					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	249	244			253					± 10
VPM MPV	µm ³ ; fl	9.0	8.6			9.0					± 0.5

ABX Minocal

LOT

CX 423

Rev 1

CAL



(Exp.) 2018-07-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 ⁹ /mm ³ ;10 ⁹ /l	8.8	8.8							± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.45	4.42							± 0.06
HB HGB	g/dl	13.6	13.5							± 0.2
	g/l	136	135							± 2
	mmol/l	8.45	8.38							± 0.12
HT HCT	%	39.2	38.7							± 1.0
	l/l	0.392	0.387							± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	232	235							± 10
VPM MPV	µm ³ ; fl	N/A	8.9							± 0.5

Ref: TEMP-0387 Rev.45 FRONT / RECTO 1300032436-A