

ABX Minocal

LOT

CX 402

CAL



(Exp.) 2016-10-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
			CRP 200			60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l		9.3			9.2					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l		4.48			4.50					± 0.06
HB HGB	g/dl		13.4			13.2					± 0.2
	g/l		134			132					± 2
	mmol/l		8.34			8.19					± 0.12
HT HCT	%		38.5			38.6					± 1.0
	l/l		0.385			0.386					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l		240			241					± 10
VPM MPV	µm ³ ; fl		9.0			8.2					± 0.5

ABX Minocal

LOT

CX 402

Rev 1

CAL



(Exp.) **2016-10-05**
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 ⁹ /mm ³ ;10 ⁹ /l	9.1	9.1							± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.52	4.52							± 0.06
HB HGB	g/dl	13.6	13.5							± 0.2
	g/l	136	135							± 2
	mmol/l	8.45	8.38							± 0.12
HT HCT	%	38.6	38.6							± 1.0
	l/l	0.386	0.386							± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	231	243							± 10
VPM MPV	µm ³ ; fl	N/A	9.0							± 0.5

Ref: TEMP-0387 Rev.42 FRONT / RECTO 1300019741-A