

ABX Minocal



LOT

CX 035

Rev 1

CAL



(Exp.) 2015-04-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.5	9.4	9.2	9.3	9.4					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.55	4.50	4.54	4.53	4.51					± 0.06
HB HGB	g/dl	13.5	13.5	13.4	13.6	13.5					± 0.2
	g/l	135	135	134	136	135					± 2
	mmol/l	8.38	8.38	8.32	8.45	8.38					± 0.12
HT HCT	%	38.4	38.5	38.7	38.7	38.6					± 1.0
	l/l	0.384	0.385	0.387	0.387	0.386					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	248	240	245	247	245					± 10
VPM MPV	µm ³ ; fl	9.1	8.7	N/A	9.4	8.3					± 0.5

Ref: TEMP-0387 Rev.38 FRONT / RECTO 9930090 -A

LOT

CX 035

CAL



(Exp.) 2015-04-05

Rev 1

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP		MS CRP	XLR	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.5		9.3	9.3	9.5	9.3	9.5	9.4	9.3	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.50		4.46	4.53	4.56	4.52	4.56	4.54	4.53	± 0.06
HB HGB	g/dl	13.2		13.3	13.5	13.4	13.4	13.4	13.5	13.5	± 0.2
	g/l	132		133	135	134	134	134	135	135	± 2
	mmol/l	8.20		8.26	8.38	8.32	8.32	8.32	8.38	8.38	± 0.12
HT HCT	%	37.8		35.9	36.8	37.2	36.2	37.2	36.8	36.8	± 1.0
	l/l	0.378		0.359	0.368	0.372	0.362	0.372	0.368	0.368	± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	240		236	242	246	245	246	245	242	± 10
VPM MPV	µm ³ ; fl	8.7		N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.38 BACK / VERSO 9930090 -A

LOT

CX 035

CAL



(Exp.) 2015-04-05

Rev 1

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
			CRP 200			60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l		9.4			9.2					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l		4.50			4.51					± 0.06
HB HGB	g/dl		13.2			13.1					± 0.2
	g/l		132			131					± 2
	mmol/l		8.22			8.13					± 0.12
HT HCT	%		38.5			38.6					± 1.0
	l/l		0.385			0.386					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l		240			245					± 10
VPM MPV	µm ³ ; fl		8.7			8.3					± 0.5

Ref: TEMP-0387 Rev.38 9930094 -A