

# ABX Minocal



**LOT** CX 094  
Rev 1

**CAL**

 (Exp.) 2014-10-05  
(YYYY - MM - DD)

| PARAMETRES<br>PARAMETERS | UNITES<br>UNITS                                       | ABX Lyse        |                |        |        |                             |        |        |        | TOLERANCES<br>TOLERANCE |         |
|--------------------------|---|-----------------|----------------|--------|--------|-----------------------------|--------|--------|--------|-------------------------|---------|
|                          |   | MICROS          | MICROS         | MICROS | MICROS | MICROS                      | PENTRA | PENTRA | PENTRA |                         | PENTRA  |
|                          |   | ES60<br>Care ST | CRP<br>CRP 200 | 8P     | 18P    | 60 8P<br>60 18P<br>ADVIA 60 |        |        |        |                         |         |
| GB WBC                   | 10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l  | 9.2             | 9.2            | 9.1    | 9.0    | 9.2                         |        |        |        |                         | ± 0.2   |
| GR RBC                   | 10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l | 4.52            | 4.50           | 4.59   | 4.56   | 4.54                        |        |        |        |                         | ± 0.06  |
| HB HGB                   | g/dl  | 13.7            | 13.7           | 13.7   | 13.7   | 13.7                        |        |        |        |                         | ± 0.2   |
|                          | g/l   | 137             | 137            | 137    | 137    | 137                         |        |        |        |                         | ± 2     |
|                          | mmol/l  | 8.51            | 8.51           | 8.51   | 8.51   | 8.51                        |        |        |        |                         | ± 0.12  |
| HT HCT                   | %   | 38.9            | 38.9           | 39.5   | 39.0   | 39.3                        |        |        |        |                         | ± 1.0   |
|                          | l/l   | 0.389           | 0.389          | 0.395  | 0.390  | 0.393                       |        |        |        |                         | ± 0.010 |
| PLAQ. PLTS               | 10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l  | 250             | 242            | 245    | 249    | 242                         |        |        |        |                         | ± 10    |
| VPM MPV                  | µm <sup>3</sup> ; fl                                  | 8.5             | 8.2            | N/A    | 8.8    | 7.9                         |        |        |        |                         | ± 0.5   |

Ref: TEMP-0387 Rev.37 FRONT / RECTO 9930090 -A

**LOT** CX 094  
Rev 1

**CAL**

 (Exp.) 2014-10-05  
(YYYY - MM - DD)

| PARAMETRES<br>PARAMETERS | UNITES<br>UNITS                                       | ABX Lysebio |        |        |        |                      |        |                                  |                     |            | TOLERANCES<br>TOLERANCE |
|--------------------------|---|-------------|--------|--------|--------|----------------------|--------|----------------------------------|---------------------|------------|-------------------------|
|                          |   | MICROS      | MICROS | PENTRA | PENTRA | PENTRA               | PENTRA | PENTRA                           | PENTRA              | PENTRA     |                         |
|                          |   | EMI CRP     |        | MS CRP | XLR    | DX NEXUS<br>DF NEXUS | MS60   | 120 - 120 RETIC<br>DX120 - DF120 | 60<br>60 C+<br>ES60 | 80<br>XL80 |                         |
| GB WBC                   | 10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l  | 9.1         |        | 8.9    | 9.1    | 9.3                  | 9.0    | 9.3                              | 9.1                 | 9.1        | ± 0.2                   |
| GR RBC                   | 10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l | 4.50        |        | 4.46   | 4.55   | 4.56                 | 4.54   | 4.56                             | 4.57                | 4.55       | ± 0.06                  |
| HB HGB                   | g/dl  | 13.5        |        | 13.5   | 13.7   | 13.6                 | 13.7   | 13.6                             | 13.6                | 13.7       | ± 0.2                   |
|                          | g/l   | 135         |        | 135    | 137    | 136                  | 137    | 136                              | 136                 | 137        | ± 2                     |
|                          | mmol/l  | 8.38        |        | 8.38   | 8.51   | 8.45                 | 8.51   | 8.45                             | 8.45                | 8.51       | ± 0.12                  |
| HT HCT                   | %   | 38.3        |        | 37.0   | 37.5   | 37.7                 | 36.8   | 37.7                             | 37.5                | 37.5       | ± 1.0                   |
|                          | l/l   | 0.383       |        | 0.370  | 0.375  | 0.377                | 0.368  | 0.377                            | 0.375               | 0.375      | ± 0.010                 |
| PLAQ. PLTS               | 10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l  | 243         |        | 240    | 242    | 246                  | 239    | 246                              | 240                 | 242        | ± 10                    |
| VPM MPV                  | µm <sup>3</sup> ; fl                                  | 8.5         |        | N/A    | N/A    | N/A                  | N/A    | N/A                              | N/A                 | N/A        | ± 0.5                   |

Ref: TEMP-0387 Rev.37 BACK / VERSO 9930090 -A

**LOT**

**CX 094**

**CAL**



**(Exp.) 2014-10-05**

(YYYY - MM - DD)

Rev 1

| PARAMETRES<br>PARAMETERS | UNITES<br>UNITS                                       | Minilysebio  |              |        |        |                 |        |        |        | TOLERANCES<br>TOLERANCE |         |
|--------------------------|---|--------------|--------------|--------|--------|-----------------|--------|--------|--------|-------------------------|---------|
|                          |   | MICROS       | MICROS       | MICROS | MICROS | MICROS          | PENTRA | PENTRA | PENTRA |                         | PENTRA  |
|                          |   | EMI CRP      | CRP 200      |        |        | 60 8P<br>60 18P |        |        |        |                         |         |
| GB WBC                   | 10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l  | <b>9.1</b>   | <b>9.2</b>   |        |        | <b>9.0</b>      |        |        |        |                         | ± 0.2   |
| GR RBC                   | 10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l | <b>4.50</b>  | <b>4.50</b>  |        |        | <b>4.54</b>     |        |        |        |                         | ± 0.06  |
| HB HGB                   | g/dl  | <b>13.5</b>  | <b>13.4</b>  |        |        | <b>13.3</b>     |        |        |        |                         | ± 0.2   |
|                          | g/l   | <b>135</b>   | <b>134</b>   |        |        | <b>133</b>      |        |        |        |                         | ± 2     |
|                          | mmol/l  | <b>8.38</b>  | <b>8.34</b>  |        |        | <b>8.25</b>     |        |        |        |                         | ± 0.12  |
| HT HCT                   | %   | <b>38.3</b>  | <b>38.9</b>  |        |        | <b>39.3</b>     |        |        |        |                         | ± 1.0   |
|                          | l/l   | <b>0.383</b> | <b>0.389</b> |        |        | <b>0.393</b>    |        |        |        |                         | ± 0.010 |
| PLAQ. PLTS               | 10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l  | <b>243</b>   | <b>242</b>   |        |        | <b>242</b>      |        |        |        |                         | ± 10    |
| VPM MPV                  | µm <sup>3</sup> ; fl                                  | <b>8.5</b>   | <b>8.2</b>   |        |        | <b>7.9</b>      |        |        |        |                         | ± 0.5   |

Ref: TEMP-0387 Rev.37 9930088-B