

LOT CX 064
Rev 1

CAL

 (Exp.) 2014-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.5	9.3	9.3	9.3	9.5					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.58	4.60	4.59	4.57	4.61					± 0.06
HB HGB	g/dl	13.8	13.8	13.7	13.8	13.9					± 0.2
	g/l	138	138	137	138	139					± 2
	mmol/l	8.57	8.57	8.51	8.57	8.63					± 0.12
HT HCT	%	39.4	39.6	39.7	39.3	39.6					± 1.0
	l/l	0.394	0.396	0.397	0.393	0.396					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	246	243	246	245	243					± 10
VPM MPV	µm ³ ; fl	8.3	8.1	N/A	8.5	7.7					± 0.5

Ref: TEMP-0387 Rev.36 FRONT / RECTO 9930081 -C

LOT CX 064
Rev 1

CAL

 (Exp.) 2014-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA		
		EMI CRP			XLR	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60		80 XL80
GB WBC	10 ⁹ /mm ³ ;10 ⁹ /l	9.4			9.3	9.6	9.3	9.6	9.4	9.3	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.57			4.57	4.62	4.56	4.62	4.58	4.57	± 0.06
HB HGB	g/dl	13.7			13.8	13.8	13.8	13.8	13.7	13.8	± 0.2
	g/l	137			138	138	138	138	137	138	± 2
	mmol/l	8.51			8.57	8.57	8.57	8.57	8.51	8.57	± 0.12
HT HCT	%	38.8			37.9	38.1	36.8	38.1	37.5	37.9	± 1.0
	l/l	0.388			0.379	0.381	0.368	0.381	0.375	0.379	± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	245			244	250	244	250	246	244	± 10
VPM MPV	µm ³ ; fl	8.4			N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.36 BACK / VERSO 9930081 -C

LOT CX 064
Rev 1

CAL

 (Exp.) 2014-07-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		EMI CRP	CRP 200			60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.4	9.3			9.3					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.57	4.60			4.61					± 0.06
HB HGB	g/dl	13.7	13.5			13.5					± 0.2
	g/l	137	135			135					± 2
	mmol/l	8.51	8.40			8.37					± 0.12
HT HCT	%	38.8	39.6			39.6					± 1.0
	l/l	0.388	0.396			0.396					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	245	243			243					± 10
VPM MPV	µm ³ ; fl	8.4	8.1			7.7					± 0.5

Ref: TEMP-0387 Rev.36 9930088-B