

ABX Minocal



LOT

CX 054

CAL



(Exp.) 2014-06-05

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.4	9.2	9.3	9.2	9.3					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.48	4.47	4.52	4.46	4.50					± 0.06
HB HGB	g/dl	13.7	13.7	13.6	13.7	13.7					± 0.2
	g/l	137	137	136	137	137					± 2
	mmol/l	8.51	8.51	8.45	8.51	8.51					± 0.12
HT HCT	%	39.0	38.9	39.3	39.0	39.2					± 1.0
	l/l	0.390	0.389	0.393	0.390	0.392					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	254	250	252	248	253					± 10
VPM MPV	µm ³ ; fl	8.4	8.2	N/A	8.4	7.9					± 0.5

Ref: TEMP-0387 Rev.36 FRONT / RECTO 9930081 -C

LOT CX 054
Rev 1

CAL

 (Exp.) 2014-06-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA		
		EMI CRP			XLR	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60		80 XL80
GB WBC	10 ⁹ /mm ³ ;10 ⁹ /l	9.3			9.3	9.5	9.3	9.5	9.4	9.3	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.47			4.47	4.50	4.43	4.50	4.47	4.47	± 0.06
HB HGB	g/dl	13.6			13.7	13.6	13.6	13.6	13.6	13.7	± 0.2
	g/l	136			137	136	136	136	136	137	± 2
	mmol/l	8.45			8.51	8.45	8.45	8.45	8.45	8.51	± 0.12
HT HCT	%	38.4			37.5	37.6	36.1	37.6	37.3	37.5	± 1.0
	l/l	0.384			0.375	0.376	0.361	0.376	0.373	0.375	± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	250			250	252	248	252	253	250	± 10
VPM MPV	µm ³ ; fl	8.5			N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.36 BACK / VERSO 9930081 -C

LOT CX 054
Rev 1

CAL

 (Exp.) 2014-06-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		EMI CRP	CRP 200			60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.3	9.2			9.1					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.47	4.47			4.50					± 0.06
HB HGB	g/dl	13.6	13.4			13.3					± 0.2
	g/l	136	134			133					± 2
	mmol/l	8.45	8.34			8.25					± 0.12
HT HCT	%	38.4	38.9			39.2					± 1.0
	l/l	0.384	0.389			0.392					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	250	250			253					± 10
VPM MPV	µm ³ ; fl	8.5	8.2			7.9					± 0.5

Ref: TEMP-0387 Rev.36 9930088-B