

ABX Minocal

LOT

CX 401

CAL



(Exp.) 2016-09-05
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|-------------|--------------|--------|--------|-----------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | | CRP 200 | | | 60 8P 60 18P | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | | 9.1 | | | 9.0 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | | 4.57 | | | 4.60 | | | | | ± 0.06 |
| HB HGB | g/dl | | 13.5 | | | 13.3 | | | | | ± 0.2 |
| | g/l | | 135 | | | 133 | | | | | ± 2 |
| | mmol/l | | 8.40 | | | 8.25 | | | | | ± 0.12 |
| HT HCT | % | | 38.8 | | | 38.9 | | | | | ± 1.0 |
| | l/l | | 0.388 | | | 0.389 | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | | 248 | | | 252 | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | | 7.6 | | | 7.2 | | | | | ± 0.5 |

ABX Minocal

LOT

CX 401
Rev 1

CAL



(Exp.) 2016-09-05
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | LyseVet | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|-----------------------------|--------|--|--|--|--|--|-------------------------|
| | | MICROS | | MICROS | | | | | | |
| | | Scil Vet ABC | Scil Vet ABC Plus ESV 60 | | | | | | | |
| GB WBC | 10 ⁹ /mm ³ ;10 ⁹ /l | 8.6 | 8.8 | | | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.56 | 4.59 | | | | | | | ± 0.06 |
| HB HGB | g/dl | 13.6 | 13.6 | | | | | | | ± 0.2 |
| | g/l | 136 | 136 | | | | | | | ± 2 |
| | mmol/l | 8.45 | 8.45 | | | | | | | ± 0.12 |
| HT HCT | % | 38.8 | 38.8 | | | | | | | ± 1.0 |
| | l/l | 0.388 | 0.388 | | | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 235 | 252 | | | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | N/A | 7.8 | | | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.42 FRONT / RECTO 1300019741-A