

ABX Minocal

LOT

CX 420

CAL



(Exp.) 2018-04-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	8.8	8.9			9.0					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.60	4.55			4.58					± 0.06
HB HGB	g/dl	14.1	13.6			14.1					± 0.2
	g/l	141	136			141					± 2
	mmol/l	8.76	8.46			8.76					± 0.12
HT HCT	%	39.7	39.4			39.6					± 1.0
	l/l	0.397	0.394			0.396					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	248	240			249					± 10
VPM MPV	µm ³ ; fl	8.7	8.4			8.6					± 0.5

ABX Minocal

LOT

CX 420
Rev 1

CAL



(Exp.) 2018-04-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 ⁹ /mm ³ ;10 ⁹ /l	8.5	8.6							± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.59	4.60							± 0.06
HB HGB	g/dl	13.8	13.7							± 0.2
	g/l	138	137							± 2
	mmol/l	8.57	8.51							± 0.12
HT HCT	%	39.7	39.7							± 1.0
	l/l	0.397	0.397							± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	235	237							± 10
VPM MPV	µm ³ ; fl	N/A	8.6							± 0.5

Ref: TEMP-0387 Rev.45 FRONT / RECTO 1300032436-A