

ABX Minocal

LOT

CX 410

CAL



(Exp.) 2017-06-05

Rev 1

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.2	9.2			9.2					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.58	4.57			4.56					± 0.06
HB HGB	g/dl	13.8	13.4			13.8					± 0.2
	g/l	138	134			138					± 2
	mmol/l	8.57	8.34			8.57					± 0.12
HT HCT	%	39.4	39.3			39.0					± 1.0
	l/l	0.394	0.393			0.390					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	254	260			257					± 10
VPM MPV	µm ³ ; fl	8.6	8.8			8.5					± 0.5

ABX Minocal

LOT

CX 410

CAL



(Exp.) 2017-06-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 ⁹ /mm ³ ;10 ⁹ /l	9.0	9.0							± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.62	4.58							± 0.06
HB HGB	g/dl	13.5	13.6							± 0.2
	g/l	135	136							± 2
	mmol/l	8.38	8.45							± 0.12
HT HCT	%	39.3	38.9							± 1.0
	l/l	0.393	0.389							± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	245	252							± 10
VPM MPV	µm ³ ; fl	N/A	8.6							± 0.5

Ref: TEMP-0387 Rev.44 FRONT / RECTO 1300032436-A