

ABX Minocal

LOT

CX 447

CAL



(Exp.) 2020-07-05

Rev 1

(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|--------------|--------------|--------|--------|-----------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 | CRP 200 | | | 60 8P 60 18P | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.3 | 9.2 | | | 9.3 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.50 | 4.46 | | | 4.45 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.5 | 13.1 | | | 13.4 | | | | | ± 0.2 |
| | g/l | 135 | 131 | | | 134 | | | | | ± 2 |
| | mmol/l | 8.38 | 8.15 | | | 8.32 | | | | | ± 0.12 |
| HT HCT | % | 38.3 | 37.9 | | | 38.5 | | | | | ± 1.0 |
| | l/l | 0.383 | 0.379 | | | 0.385 | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 245 | 247 | | | 245 | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | 8.2 | 7.7 | | | 8.8 | | | | | ± 0.5 |

ABX Minocal

LOT

CX 447

CAL



(Exp.) 2020-07-05
(YYYY-MM-DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | LyseVet | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|-----------------------------|--------|--|--|--|--|--|-------------------------|
| | | MICROS | | MICROS | | | | | | |
| | | Scil Vet ABC | Scil Vet ABC Plus ESV 60 | | | | | | | |
| GB WBC | 10 ⁹ /mm ³ ;10 ⁹ /l | 9.0 | 9.0 | | | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.46 | 4.45 | | | | | | | ± 0.06 |
| HB HGB | g/dl | 13.2 | 13.1 | | | | | | | ± 0.2 |
| | g/l | 132 | 131 | | | | | | | ± 2 |
| | mmol/l | 8.20 | 8.14 | | | | | | | ± 0.12 |
| HT HCT | % | 37.9 | 37.8 | | | | | | | ± 1.0 |
| | l/l | 0.379 | 0.378 | | | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 235 | 246 | | | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | N/A | 8.6 | | | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.46 FRONT / RECTO 1300060881