

LOT MX 405 L
 Rev 1

CONTROL L

(Exp.)

2017-08-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	1.9	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.41	2.40			2.35					± 0.15
HB HGB	g/dl	6.0	6.1			6.0					± 0.4
	g/l	60	61			60					± 4
	mmol/l	3.73	3.77			3.73					± 0.25
HT HCT	%	17.1	16.8			16.5					± 2.0
	l/l	0.171	0.168			0.165					± 0.020
VGM MCV	µm ³ ; fl	71	70			70					± 4
TGMH MCH	pg	24.9	25.3			25.5					± 2.0
	fmol	1.55	1.57			1.59					± 0.12
CCMH MCHC	g/dl	35.1	36.2			36.5					± 3.0
	g/l	351	362			365					± 30
	mmol/l	21.78	22.46			22.65					± 1.86
IDR RDW	%	14.7	14.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	68	75			75					± 20
VPM MPV	µm ³ ; fl	9.0	8.8			8.6					± 2.0
LYMPHO	#	1.27	1.31			1.30					± 0.40
	%	66.7	65.7			65.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.1	10.0			8.0					± 6.0
GRANULO	#	0.48	0.49			0.54					± 0.40
	%	25.2	24.5			27.0					± 7.0

LOT **MX 405** **N**
 Rev 1

CONTROL **N**

(Exp.)

2017-08-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 BP - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.7			7.8					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.89	4.83			4.78					± 0.18
HB HGB	g/dl	13.7	13.7			13.8					± 0.5
	g/l	137	137			138					± 5
	mmol/l	8.51	8.51			8.57					± 0.31
HT HCT	%	40.1	38.6			38.2					± 2.5
	l/l	0.401	0.386			0.382					± 0.025
VGM MCV	µm ³ ; fl	82	80			80					± 4
TGMH MCH	pg	28.0	28.4			28.9					± 2.0
	fmol	1.74	1.76			1.79					± 0.12
CCMH MCHC	g/dl	34.2	35.5			36.1					± 3.0
	g/l	342	355			361					± 30
	mmol/l	21.22	22.02			22.41					± 1.86
IDR RDW	%	15.4	13.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	256	265			258					± 40
VPM MPV	µm ³ ; fl	8.6	8.5			8.1					± 2.0
LYMPHO	#	2.63	2.65			2.69					± 0.40
	%	34.2	34.5			34.5					± 5.0
MONO	#	0.59	0.65			0.59					± 0.30
	%	7.7	8.5			7.5					± 4.0
GRANULO	#	4.47	4.38			4.52					± 0.60
	%	58.1	56.8			58.0					± 6.0

1300027266-A
 Ref: TEMP-0830 Rev.36

LOT MX 405 H
Rev 1

CONTROL H

 (Exp.)

2017-08-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	19.9	20.3			20.1					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.78	5.70			5.71					± 0.20
HB HGB	g/dl	18.3	17.9			18.5					± 0.6
	g/l	183	179			185					± 6
	mmol/l	11.36	11.12			11.49					± 0.37
HT HCT	%	52.6	50.2			50.8					± 3.0
	l/l	0.526	0.502			0.508					± 0.030
VGM MCV	µm ³ ; fl	91	88			89					± 4
TGMH MCH	pg	31.7	31.4			32.4					± 2.0
	fmol	1.97	1.95			2.01					± 0.12
CCMH MCHC	g/dl	34.8	35.7			36.4					± 3.0
	g/l	348	357			364					± 30
	mmol/l	21.61	22.16			22.61					± 1.86
IDR RDW	%	15.0	13.0			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	509	510			505					± 45
VPM MPV	µm ³ ; fl	8.1	8.0			7.7					± 2.0
LYMPHO	#	3.10	3.23			3.32					± 1.00
	%	15.6	15.9			16.5					± 5.0
MONO	#	1.23	1.53			1.01					± 0.80
	%	6.2	7.6			5.0					± 4.0
GRANULO	#	15.56	15.52			15.78					± 1.40
	%	78.2	76.4			78.5					± 7.0