

LOT MX 422 L
 Rev 1

CONTROL L

(Exp.)

2020-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.34	2.35			2.33					± 0.15
HB HGB	g/dl	5.9	5.8			5.9					± 0.4
	g/l	59	58			59					± 4
	mmol/l	3.66	3.59			3.66					± 0.25
HT HCT	%	16.6	16.5			16.5					± 2.0
	l/l	0.166	0.165			0.165					± 0.020
VGM MCV	µm ³ ; fl	71	70			71					± 4
TGMH MCH	pg	25.2	24.6			25.3					± 2.0
	fmol	1.57	1.53			1.57					± 0.12
CCMH MCHC	g/dl	35.5	35.1			35.7					± 3.0
	g/l	355	351			357					± 30
	mmol/l	22.05	21.83			22.15					± 1.86
IDR RDW	%	14.2	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74	73			76					± 20
VPM MPV	µm ³ ; fl	8.6	8.7			9.0					± 2.0
LYMPHO	#	1.31	1.31			1.31					± 0.40
	%	65.5	65.7			65.5					± 8.0
MONO	#	0.20	0.22			0.20					± 0.20
	%	9.5	11.0			9.5					± 6.0
GRANULO	#	0.50	0.47			0.50					± 0.40
	%	25.0	23.5			25.0					± 7.0

LOT MX 422 N
 Rev 1

CONTROL N

(Exp.)

2020-06-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.5			7.8					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.67	4.68			4.69					± 0.18
HB HGB	g/dl	13.3	13.2			13.5					± 0.5
	g/l	133	132			135					± 5
	mmol/l	8.26	8.20			8.38					± 0.31
HT HCT	%	36.9	37.0			37.5					± 2.5
	l/l	0.369	0.370			0.375					± 0.025
VGM MCV	µm ³ ; fl	79	79			80					± 4
TGMH MCH	pg	28.5	28.2			28.8					± 2.0
	fmol	1.77	1.75			1.79					± 0.12
CCMH MCHC	g/dl	36.1	35.7			36.0					± 3.0
	g/l	361	357			360					± 30
	mmol/l	22.39	22.17			22.34					± 1.86
IDR RDW	%	14.0	13.5			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	260	260			260					± 40
VPM MPV	µm ³ ; fl	7.9	7.8			8.2					± 2.0
LYMPHO	#	2.70	2.82			2.69					± 0.40
	%	35.0	37.6			34.5					± 5.0
MONO	#	0.50	0.43			0.47					± 0.30
	%	6.5	5.8			6.0					± 4.0
GRANULO	#	4.50	4.26			4.64					± 0.60
	%	58.5	56.8			59.5					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626



LOT		MX 422 H	CONTROL H				(Exp.)				2020-06-05 (YYYY-MM-DD)
Rev 1		Minilysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	21.1	20.3			21.2					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.69	5.67			5.69					± 0.20
HB HGB	g/dl	18.6	17.9			18.7					± 0.6
	g/l	186	179			187					± 6
	mmol/l	11.55	11.12			11.61					± 0.37
HT HCT	%	50.1	49.9			50.6					± 3.0
	l/l	0.501	0.499			0.506					± 0.030
VGM MCV	µm ³ ; fl	88	88			89					± 4
TGMH MCH	pg	32.7	31.6			32.9					± 2.0
	fmol	2.03	1.96			2.04					± 0.12
CCMH MCHC	g/dl	37.1	35.9			36.9					± 3.0
	g/l	371	359			369					± 30
	mmol/l	23.07	22.28			22.93					± 1.86
IDR RDW	%	13.0	12.5			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	510	510			520					± 45
VPM MPV	µm ³ ; fl	8.3	8.2			8.6					± 2.0
LYMPHO	#	3.69	3.55			3.50					± 1.00
	%	17.5	17.5			16.5					± 5.0
MONO	#	0.84	0.99			0.85					± 0.80
	%	4.0	4.9			4.0					± 4.0
GRANULO	#	16.56	15.72			16.85					± 1.40
	%	78.5	77.4			79.5					± 7.0