

**LOT** MX 432 L  
Rev 1

**CONTROL** L

 (Exp.)

**2022-02-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.0			2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.30	2.31			2.27					± 0.17
HB HGB	g/dl	6.0	5.9			5.7					± 0.5
	g/l	60	59			57					± 5
	mmol/l	3.73	3.65			3.52					± 0.31
HT HCT	%	16.3	16.2			16.6					± 2.0
	l/l	0.163	0.162			0.166					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	71	70			73					± 4
TGMH MCH	pg	26.1	25.5			25.0					± 2.0
	fmol	1.62	1.58			1.55					± 0.12
CCMH MCHC	g/dl	36.7	36.4			34.2					± 3.0
	g/l	367	364			342					± 30
	mmol/l	22.82	22.58			21.26					± 1.86
IDR RDW	%	15.0	14.0			15.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	80			78					± 22
VPM MPV	µm <sup>3</sup> ; fl	8.4	8.3			7.6					± 2.0
LYMPHO	#	1.29	1.29			1.30					± 0.40
	%	64.5	64.6			65.0					± 8.0
MONO	#	0.20	0.26			0.20					± 0.20
	%	10.0	13.0			9.5					± 6.0
GRANULO	#	0.51	0.45			0.51					± 0.40
	%	25.5	22.5			25.5					± 7.0

**LOT** MX 432 N  
 Rev 1


**CONTROL** N

(Exp.)

**2022-02-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.5	7.3			7.3					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.55	4.53			4.53					± 0.26
HB HGB	g/dl	13.4	13.4			13.2					± 0.7
	g/l	134	134			132					± 7
	mmol/l	8.32	8.32			8.22					± 0.43
HT HCT	%	37.3	37.1			37.6					± 2.7
	l/l	0.373	0.371			0.376					± 0.027
VGM MCV	µm <sup>3</sup> ; fl	82	82			83					± 4
TGMH MCH	pg	29.5	29.6			29.2					± 2.0
	fmol	1.83	1.84			1.81					± 0.12
CCMH MCHC	g/dl	35.9	36.1			35.2					± 3.0
	g/l	359	361			352					± 30
	mmol/l	22.30	22.40			21.85					± 1.86
IDR RDW	%	14.0	13.5			13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	238	250			250					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.4	7.7			7.4					± 2.0
LYMPHO	#	2.74	2.79			2.63					± 0.40
	%	36.5	38.2			36.0					± 5.0
MONO	#	0.38	0.45			0.33					± 0.30
	%	5.0	6.2			4.5					± 4.0
GRANULO	#	4.39	4.08			4.34					± 0.60
	%	58.5	55.9			59.5					± 6.0

Ref: TEMP-0830 Rev.41 FRONT / RECTO 130060626

LOT		MX 432 H		CONTROL		H		 (Exp.) <span style="float: right;">2022-02-05 (YYYY-MM-DD)</span>			
Rev 1		Minilysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	20.0	19.7			19.6					± 1.7
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.64	5.63			5.65					± 0.32
HB HGB	g/dl	18.4	18.1			18.0					± 0.9
	g/l	184	181			180					± 9
	mmol/l	11.43	11.24			11.19					± 0.56
HT HCT	%	50.2	50.1			51.4					± 3.4
	l/l	0.502	0.501			0.514					± 0.034
VGM MCV	µm <sup>3</sup> ; fl	89	89			91					± 4
TGMH MCH	pg	32.6	32.1			31.9					± 2.0
	fmol	2.03	2.00			1.98					± 0.12
CCMH MCHC	g/dl	36.7	36.1			35.0					± 3.0
	g/l	367	361			350					± 30
	mmol/l	22.76	22.43			21.76					± 1.86
IDR RDW	%	14.5	14.0			14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	472	485			495					± 65
VPM MPV	µm <sup>3</sup> ; fl	8.0	7.7			7.3					± 2.0
LYMPHO	#	3.20	2.92			3.04					± 1.00
	%	16.0	14.8			15.5					± 5.0
MONO	#	1.20	1.49			1.08					± 0.80
	%	6.0	7.6			5.5					± 4.0
GRANULO	#	15.60	15.25			15.48					± 1.40
	%	78.0	77.4			79.0					± 7.0