

LOT **MX 403 L**
Rev 1

CONTROL **L**

 (Exp.)

2017-04-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	1.9	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.40			2.39					± 0.15
HB HGB	g/dl	6.0	6.1			6.0					± 0.4
	g/l	60	61			60					± 4
	mmol/l	3.73	3.77			3.73					± 0.25
HT HCT	%	16.6	16.6			16.7					± 2.0
	l/l	0.166	0.166			0.167					± 0.020
VGM MCV	µm ³ ; fl	69	69			70					± 4
TGMH MCH	pg	25.0	25.3			25.1					± 2.0
	fmol	1.55	1.57			1.56					± 0.12
CCMH MCHC	g/dl	36.2	36.7			35.9					± 3.0
	g/l	362	367			359					± 30
	mmol/l	22.50	22.79			22.27					± 1.86
IDR RDW	%	14.0	13.5			13.8					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	73	70			74					± 20
VPM MPV	µm ³ ; fl	9.0	9.0			8.9					± 2.0
LYMPHO	#	1.25	1.30			1.32					± 0.40
	%	66.0	65.1			66.0					± 8.0
MONO	#	0.20	0.19			0.20					± 0.20
	%	7.5	9.5			8.0					± 6.0
GRANULO	#	0.50	0.51			0.52					± 0.40
	%	26.5	25.5			26.0					± 7.0

LOT MX 403 N
Rev 1

CONTROL N

 (Exp.)

2017-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.4	7.6			7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.70	4.68			4.67					± 0.18
HB HGB	g/dl	13.6	13.6			13.6					± 0.5
	g/l	136	136			136					± 5
	mmol/l	8.45	8.45			8.45					± 0.31
HT HCT	%	37.6	37.4			37.4					± 2.5
	l/l	0.376	0.374			0.374					± 0.025
VGM MCV	µm ³ ; fl	80	80			80					± 4
TGMH MCH	pg	28.9	29.1			29.1					± 2.0
	fmol	1.80	1.80			1.81					± 0.12
CCMH MCHC	g/dl	36.2	36.3			36.4					± 3.0
	g/l	362	363			364					± 30
	mmol/l	22.46	22.56			22.61					± 1.86
IDR RDW	%	13.8	13.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	247	252			249					± 40
VPM MPV	µm ³ ; fl	8.6	8.8			8.5					± 2.0
LYMPHO	#	2.65	2.86			2.77					± 0.40
	%	35.8	37.6			36.0					± 5.0
MONO	#	0.38	0.44			0.42					± 0.30
	%	5.2	5.8			5.5					± 4.0
GRANULO	#	4.37	4.32			4.50					± 0.60
	%	59.0	56.8			58.5					± 6.0

1300027266-A
Ref: TEMP-0830 Rev.36

LOT MX 403 H
Rev 1

CONTROL H

 (Exp.)

2017-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.0	20.0			20.0					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.68	5.62			5.60					± 0.20
HB HGB	g/dl	18.0	17.8			18.1					± 0.6
	g/l	180	178			181					± 6
	mmol/l	11.18	11.05			11.24					± 0.37
HT HCT	%	50.0	49.5			49.3					± 3.0
	l/l	0.500	0.495			0.493					± 0.030
VGM MCV	µm ³ ; fl	88	88			88					± 4
TGMH MCH	pg	31.7	31.7			32.3					± 2.0
	fmol	1.97	1.97			2.01					± 0.12
CCMH MCHC	g/dl	36.0	36.0			36.7					± 3.0
	g/l	360	360			367					± 30
	mmol/l	22.36	22.35			22.81					± 1.86
IDR RDW	%	13.5	12.5			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	505	515			505					± 45
VPM MPV	µm ³ ; fl	8.6	8.7			8.4					± 2.0
LYMPHO	#	3.36	3.18			3.30					± 1.00
	%	16.8	15.9			16.5					± 5.0
MONO	#	1.00	1.08			0.90					± 0.80
	%	5.0	5.4			4.5					± 4.0
GRANULO	#	15.64	15.68			15.80					± 1.40
	%	78.2	78.4			79.0					± 7.0