



LOT MX 413 L
 Rev 1

CONTROL L



2018-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1			2.2					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.31	2.29			2.30					± 0.15
HB HGB	g/dl	5.8	5.8			5.7					± 0.4
	g/l	58	58			57					± 4
	mmol/l	3.60	3.59			3.54					± 0.25
HT HCT	%	16.2	15.8			16.1					± 2.0
	l/l	0.162	0.158			0.161					± 0.020
VGM MCV	µm ³ ; fl	70	69			70					± 4
TGMH MCH	pg	25.1	25.2			24.8					± 2.0
	fmol	1.56	1.57			1.54					± 0.12
CCMH MCHC	g/dl	35.9	36.6			35.4					± 3.0
	g/l	359	366			354					± 30
	mmol/l	22.27	22.72			21.99					± 1.86
IDR RDW	%	14.5	13.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	78	80			84					± 20
VPM MPV	µm ³ ; fl	9.5	8.8			8.8					± 2.0
LYMPHO	#	1.42	1.42			1.47					± 0.40
	%	67.5	67.7			67.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	7.5	9.5			7.5					± 6.0
GRANULO	#	0.53	0.48			0.56					± 0.40
	%	25.0	23.0			25.5					± 7.0

LOT MX 413 N
 Rev 1

CONTROL N

(Exp.)

2018-12-05
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.5	7.5			7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.59	4.54			4.58					± 0.18
HB HGB	g/dl	13.5	13.4			13.5					± 0.5
	g/l	135	134			135					± 5
	mmol/l	8.38	8.32			8.38					± 0.31
HT HCT	%	36.7	36.8			37.1					± 2.5
	l/l	0.367	0.368			0.371					± 0.025
VGM MCV	µm ³ ; fl	80	81			81					± 4
TGMH MCH	pg	29.4	29.5			29.5					± 2.0
	fmol	1.83	1.83			1.83					± 0.12
CCMH MCHC	g/dl	36.8	36.4			36.4					± 3.0
	g/l	368	364			364					± 30
	mmol/l	22.83	22.63			22.60					± 1.86
IDR RDW	%	14.5	13.5			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	262	262			268					± 40
VPM MPV	µm ³ ; fl	8.9	8.4			8.5					± 2.0
LYMPHO	#	2.74	2.86			2.77					± 0.40
	%	36.5	38.2			36.0					± 5.0
MONO	#	0.38	0.40			0.42					± 0.30
	%	5.0	5.3			5.5					± 4.0
GRANULO	#	4.39	4.26			4.50					± 0.60
	%	58.5	56.8			58.5					± 6.0

1300027266-A
 Ref: TEMP-0830 Rev.36

LOT

MX 413 H
Rev 1

CONTROL H

 (Exp.)

2018-12-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.3	20.3			20.7					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.62	5.56			5.61					± 0.20
HB HGB	g/dl	18.3	17.5			18.3					± 0.6
	g/l	183	175			183					± 6
	mmol/l	11.36	10.87			11.36					± 0.37
HT HCT	%	49.5	49.5			49.9					± 3.0
	l/l	0.495	0.495			0.499					± 0.030
VGM MCV	µm ³ ; fl	88	89			89					± 4
TGMH MCH	pg	32.6	31.5			32.6					± 2.0
	fmol	2.02	1.95			2.03					± 0.12
CCMH MCHC	g/dl	37.0	35.4			36.7					± 3.0
	g/l	370	354			367					± 30
	mmol/l	22.98	21.96			22.76					± 1.86
IDR RDW	%	13.5	13.0			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	505	513			520					± 55
VPM MPV	µm ³ ; fl	8.6	8.2			8.3					± 2.0
LYMPHO	#	3.55	3.55			3.62					± 1.00
	%	17.5	17.5			17.5					± 5.0
MONO	#	1.02	1.43			1.04					± 0.80
	%	5.0	7.0			5.0					± 4.0
GRANULO	#	15.73	15.32			16.04					± 1.40
	%	77.5	75.5			77.5					± 7.0