

LOT MX 433 L
Rev 1

CONTROL L

 (Exp.)

2022-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.32	2.30			2.27					± 0.17
HB HGB	g/dl	6.0	6.1			5.8					± 0.5
	g/l	60	61			58					± 5
	mmol/l	3.73	3.77			3.58					± 0.31
HT HCT	%	16.5	16.6			16.6					± 2.0
	l/l	0.165	0.166			0.166					± 0.020
VGM MCV	µm ³ ; fl	71	72			73					± 4
TGMH MCH	pg	25.9	26.4			25.4					± 2.0
	fmol	1.61	1.64			1.58					± 0.12
CCMH MCHC	g/dl	36.4	36.7			34.8					± 3.0
	g/l	364	367			348					± 30
	mmol/l	22.62	22.79			21.61					± 1.86
IDR RDW	%	15.0	14.0			15.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	78			80					± 22
VPM MPV	µm ³ ; fl	9.2	8.5			8.5					± 2.0
LYMPHO	#	1.39	1.40			1.40					± 0.40
	%	66.0	66.7			66.5					± 8.0
MONO	#	0.17	0.22			0.16					± 0.16
	%	8.0	10.5			7.5					± 6.0
GRANULO	#	0.55	0.48			0.55					± 0.40
	%	26.0	23.0			26.0					± 7.0

LOT

MX 433 N
Rev 1

CONTROL N

(Exp.)

2022-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.6			7.6					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.49	4.47			4.44					± 0.26
HB HGB	g/dl	13.4	13.4			13.2					± 0.7
	g/l	134	134			132					± 7
	mmol/l	8.32	8.32			8.22					± 0.43
HT HCT	%	36.8	37.1			37.3					± 2.7
	l/l	0.368	0.371			0.373					± 0.027
VGM MCV	µm ³ ; fl	82	83			84					± 4
TGMH MCH	pg	29.8	30.0			29.8					± 2.0
	fmol	1.85	1.86			1.85					± 0.12
CCMH MCHC	g/dl	36.4	36.1			35.5					± 3.0
	g/l	364	361			355					± 30
	mmol/l	22.60	22.43			22.03					± 1.86
IDR RDW	%	15.0	14.0			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	255	264			267					± 40
VPM MPV	µm ³ ; fl	8.4	8.0			7.8					± 2.0
LYMPHO	#	2.70	2.78			2.62					± 0.40
	%	35.0	36.6			34.5					± 5.0
MONO	#	0.46	0.51			0.46					± 0.30
	%	6.0	6.7			6.0					± 4.0
GRANULO	#	4.54	4.32			4.52					± 0.60
	%	59.0	56.8			59.5					± 6.0

Ref: TEMP-0830 Rev.41 FRONT / RECTO 130060626

LOT MX 433 H
Rev 1

CONTROL H

 (Exp.)

2022-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.1	19.8			19.5					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.60	5.54			5.59					± 0.32
HB HGB	g/dl	18.1	17.8			17.7					± 0.9
	g/l	181	178			177					± 9
	mmol/l	11.24	11.05			11.00					± 0.56
HT HCT	%	49.3	49.3			49.8					± 3.4
	l/l	0.493	0.493			0.498					± 0.034
VGM MCV	µm ³ ; fl	88	89			89					± 4
TGMH MCH	pg	32.3	32.1			31.7					± 2.0
	fmol	2.01	2.00			1.97					± 0.12
CCMH MCHC	g/dl	36.7	36.1			35.6					± 3.0
	g/l	367	361			356					± 30
	mmol/l	22.81	22.42			22.12					± 1.86
IDR RDW	%	13.5	13.0			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	467	475			485					± 65
VPM MPV	µm ³ ; fl	8.6	8.2			8.0					± 2.0
LYMPHO	#	3.72	3.57			3.51					± 1.00
	%	18.5	18.0			18.0					± 5.0
MONO	#	0.90	1.07			0.78					± 0.78
	%	4.5	5.4			4.0					± 4.0
GRANULO	#	15.48	15.14			15.21					± 1.40
	%	77.0	76.4			78.0					± 7.0