

LOT MX 404 L
Rev 1

CONTROL L

 (Exp.)

2017-06-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.1			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.43	2.40			2.42					± 0.15
HB HGB	g/dl	6.0	6.2			6.0					± 0.4
	g/l	60	62			60					± 4
	mmol/l	3.73	3.83			3.73					± 0.25
HT HCT	%	16.5	16.6			16.5					± 2.0
	l/l	0.165	0.166			0.165					± 0.020
VGM MCV	µm ³ ; fl	68	69			68					± 4
TGMH MCH	pg	24.7	25.7			24.8					± 2.0
	fmol	1.53	1.60			1.54					± 0.12
CCMH MCHC	g/dl	36.3	37.3			36.5					± 3.0
	g/l	363	373			365					± 30
	mmol/l	22.55	23.15			22.64					± 1.86
IDR RDW	%	14.3	13.9			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	77			81					± 20
VPM MPV	µm ³ ; fl	9.2	9.0			8.6					± 2.0
LYMPHO	#	1.32	1.39			1.40					± 0.40
	%	66.0	66.2			66.5					± 8.0
MONO	#	0.20	0.19			0.20					± 0.20
	%	8.0	9.0			8.0					± 6.0
GRANULO	#	0.52	0.52			0.54					± 0.40
	%	26.0	25.0			25.5					± 7.0

LOT MX 404 N
Rev 1

CONTROL N

 (Exp.)

2017-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.4	7.5			7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.77	4.73			4.79					± 0.18
HB HGB	g/dl	13.3	13.4			13.5					± 0.5
	g/l	133	134			135					± 5
	mmol/l	8.26	8.32			8.38					± 0.31
HT HCT	%	36.7	36.4			36.9					± 2.5
	l/l	0.367	0.364			0.369					± 0.025
VGM MCV	µm ³ ; fl	77	77			77					± 4
TGMH MCH	pg	27.9	28.3			28.2					± 2.0
	fmol	1.73	1.76			1.75					± 0.12
CCMH MCHC	g/dl	36.2	36.8			36.6					± 3.0
	g/l	362	368			366					± 30
	mmol/l	22.49	22.85			22.73					± 1.86
IDR RDW	%	15.3	13.6			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	263	263			267					± 40
VPM MPV	µm ³ ; fl	9.0	8.8			8.4					± 2.0
LYMPHO	#	2.63	2.78			2.77					± 0.40
	%	35.5	37.1			36.0					± 5.0
MONO	#	0.48	0.47			0.50					± 0.30
	%	6.5	6.2			6.5					± 4.0
GRANULO	#	4.29	4.26			4.43					± 0.60
	%	58.0	56.8			57.5					± 6.0

1300027266-A
Ref: TEMP-0830 Rev.36

LOT MX 404 H
Rev 1

CONTROL H

 (Exp.)

2017-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.2	20.2			20.8					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.61	5.63			5.72					± 0.20
HB HGB	g/dl	17.9	17.8			18.3					± 0.6
	g/l	179	178			183					± 6
	mmol/l	11.12	11.05			11.36					± 0.37
HT HCT	%	48.8	49.5			49.8					± 3.0
	l/l	0.488	0.495			0.498					± 0.030
VGM MCV	µm ³ ; fl	87	88			87					± 4
TGMH MCH	pg	31.9	31.6			32.0					± 2.0
	fmol	1.98	1.96			1.99					± 0.12
CCMH MCHC	g/dl	36.7	35.9			36.8					± 3.0
	g/l	367	359			368					± 30
	mmol/l	22.78	22.31			22.84					± 1.86
IDR RDW	%	14.2	13.2			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	500	515			520					± 45
VPM MPV	µm ³ ; fl	8.2	8.0			7.8					± 2.0
LYMPHO	#	3.43	3.43			3.54					± 1.00
	%	17.0	17.0			17.0					± 5.0
MONO	#	1.01	1.20			1.04					± 0.80
	%	5.0	5.9			5.0					± 4.0
GRANULO	#	15.76	15.54			16.22					± 1.40
	%	78.0	76.9			78.0					± 7.0