

LOT MX 414 L
 Rev 1

CONTROL L

(Exp.)

2019-02-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.36	2.35			2.35					± 0.15
HB HGB	g/dl	5.9	6.0			5.8					± 0.4
	g/l	59	60			58					± 4
	mmol/l	3.66	3.71			3.60					± 0.25
HT HCT	%	16.3	16.5			16.2					± 2.0
	l/l	0.163	0.165			0.162					± 0.020
VGM MCV	µm ³ ; fl	69	70			69					± 4
TGMH MCH	pg	25.0	25.4			24.7					± 2.0
	fmol	1.55	1.58			1.53					± 0.12
CCMH MCHC	g/dl	36.2	36.3			35.8					± 3.0
	g/l	362	363			358					± 30
	mmol/l	22.50	22.57			22.21					± 1.86
IDR RDW	%	14.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	80	77			82					± 20
VPM MPV	µm ³ ; fl	9.2	8.8			9.0					± 2.0
LYMPHO	#	1.39	1.37			1.37					± 0.40
	%	66.0	65.1			65.0					± 8.0
MONO	#	0.20	0.21			0.20					± 0.20
	%	8.0	10.0			8.0					± 6.0
GRANULO	#	0.55	0.52			0.57					± 0.40
	%	26.0	25.0			27.0					± 7.0

LOT MX 414 N
 Rev 1

CONTROL N

(Exp.)

2019-02-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.6	7.5			7.8					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.62	4.60			4.61					± 0.18
HB HGB	g/dl	13.3	13.3			13.3					± 0.5
	g/l	133	133			133					± 5
	mmol/l	8.26	8.26			8.26					± 0.31
HT HCT	%	37.0	36.8			36.9					± 2.5
	l/l	0.370	0.368			0.369					± 0.025
VGM MCV	µm ³ ; fl	80	80			80					± 4
TGMH MCH	pg	28.8	28.9			28.9					± 2.0
	fmol	1.79	1.80			1.79					± 0.12
CCMH MCHC	g/dl	36.0	36.1			36.1					± 3.0
	g/l	360	361			361					± 30
	mmol/l	22.35	22.44			22.40					± 1.86
IDR RDW	%	14.0	13.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	252	257			260					± 40
VPM MPV	µm ³ ; fl	8.7	8.1			8.5					± 2.0
LYMPHO	#	2.74	2.70			2.81					± 0.40
	%	36.0	36.0			36.0					± 5.0
MONO	#	0.49	0.50			0.51					± 0.30
	%	6.5	6.7			6.5					± 4.0
GRANULO	#	4.37	4.30			4.49					± 0.60
	%	57.5	57.3			57.5					± 6.0

1300027266-A
 Ref: TEMP-0830 Rev.36

LOT MX 414 H
Rev 1

CONTROL H

 (Exp.)

2019-02-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.1	20.3			20.9					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.70	5.65			5.70					± 0.20
HB HGB	g/dl	18.3	17.8			18.3					± 0.6
	g/l	183	178			183					± 6
	mmol/l	11.36	11.05			11.36					± 0.37
HT HCT	%	50.7	50.3			50.7					± 3.0
	l/l	0.507	0.503			0.507					± 0.030
VGM MCV	µm ³ ; fl	89	89			89					± 4
TGMH MCH	pg	32.1	31.5			32.1					± 2.0
	fmol	1.99	1.96			1.99					± 0.12
CCMH MCHC	g/dl	36.1	35.4			36.1					± 3.0
	g/l	361	354			361					± 30
	mmol/l	22.40	21.98			22.40					± 1.86
IDR RDW	%	13.5	13.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	470	475			480					± 45
VPM MPV	µm ³ ; fl	8.6	8.0			8.4					± 2.0
LYMPHO	#	3.32	3.34			3.55					± 1.00
	%	16.5	16.4			17.0					± 5.0
MONO	#	1.01	1.32			1.05					± 0.80
	%	5.0	6.5			5.0					± 4.0
GRANULO	#	15.78	15.62			16.30					± 1.40
	%	78.5	76.9			78.0					± 7.0