

**LOT** MX 424 L  
Rev 1

**CONTROL** L

 (Exp.)

**2020-10-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.0			2.1					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.34			2.35					± 0.15
HB HGB	g/dl	5.9	5.8			5.8					± 0.4
	g/l	59	58			58					± 4
	mmol/l	3.66	3.59			3.60					± 0.25
HT HCT	%	16.7	16.1			16.5					± 2.0
	l/l	0.167	0.161			0.165					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	70	69			70					± 4
TGMH MCH	pg	24.8	24.7			24.7					± 2.0
	fmol	1.54	1.53			1.53					± 0.12
CCMH MCHC	g/dl	35.4	35.8			35.3					± 3.0
	g/l	354	358			353					± 30
	mmol/l	21.99	22.24			21.90					± 1.86
IDR RDW	%	14.5	14.0			14.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73	74			74					± 20
VPM MPV	µm <sup>3</sup> ; fl	8.5	8.0			8.6					± 2.0
LYMPHO	#	1.28	1.27			1.34					± 0.40
	%	64.0	63.6			64.0					± 8.0
MONO	#	0.20	0.22			0.20					± 0.20
	%	9.0	11.0			9.0					± 6.0
GRANULO	#	0.54	0.51			0.57					± 0.40
	%	27.0	25.5			27.0					± 7.0



**LOT** MX 424 N  
 Rev 1

**CONTROL** N

(Exp.)

**2020-10-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.9	7.7			7.8					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.62	4.56			4.55					± 0.18
HB HGB	g/dl	13.6	13.3			13.4					± 0.5
	g/l	136	133			134					± 5
	mmol/l	8.45	8.26			8.32					± 0.31
HT HCT	%	37.4	36.9			37.3					± 2.5
	l/l	0.374	0.369			0.373					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	81	81			82					± 4
TGMH MCH	pg	29.4	29.2			29.5					± 2.0
	fmol	1.83	1.81			1.83					± 0.12
CCMH MCHC	g/dl	36.3	36.0			35.9					± 3.0
	g/l	363	360			359					± 30
	mmol/l	22.57	22.36			22.30					± 1.86
IDR RDW	%	14.0	13.8			14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	265	261			265					± 40
VPM MPV	µm <sup>3</sup> ; fl	7.9	7.6			8.0					± 2.0
LYMPHO	#	2.77	2.78			2.73					± 0.40
	%	35.0	36.0			35.0					± 5.0
MONO	#	0.51	0.51			0.51					± 0.30
	%	6.5	6.7			6.5					± 4.0
GRANULO	#	4.62	4.41			4.56					± 0.60
	%	58.5	57.3			58.5					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626

**LOT** MX 424 H  
Rev 1

**CONTROL** H

 (Exp.)

**2020-10-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	20.8	20.1			20.6					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.72	5.62			5.65					± 0.20
HB HGB	g/dl	18.5	17.8			18.4					± 0.6
	g/l	185	178			184					± 6
	mmol/l	11.49	11.05			11.43					± 0.37
HT HCT	%	50.9	50.0			50.3					± 3.0
	l/l	0.509	0.500			0.503					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	89	89			89					± 4
TGMH MCH	pg	32.3	31.7			32.6					± 2.0
	fmol	2.01	1.97			2.02					± 0.12
CCMH MCHC	g/dl	36.3	35.6			36.6					± 3.0
	g/l	363	356			366					± 30
	mmol/l	22.57	22.10			22.72					± 1.86
IDR RDW	%	13.2	13.2			13.8					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	513	510			515					± 45
VPM MPV	µm <sup>3</sup> ; fl	7.8	7.4			7.8					± 2.0
LYMPHO	#	3.54	3.52			3.50					± 1.00
	%	17.0	17.5			17.0					± 5.0
MONO	#	0.83	1.09			0.82					± 0.80
	%	4.0	5.4			4.0					± 4.0
GRANULO	#	16.43	15.46			16.27					± 1.40
	%	79.0	76.9			79.0					± 7.0