



LOT

MX 434 L

CONTROL L



2022-06-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.24	2.26			2.20					± 0.17
HB HGB	g/dl	5.9	6.0			5.6					± 0.5
	g/l	59	60			56					± 5
	mmol/l	3.66	3.71			3.47					± 0.31
HT HCT	%	16.1	16.3			16.1					± 2.0
	l/l	0.161	0.163			0.161					± 0.020
VGM MCV	µm ³ ; fl	72	72			73					± 4
TGMH MCH	pg	26.3	26.5			25.4					± 2.0
	fmol	1.64	1.64			1.58					± 0.12
CCMH MCHC	g/dl	36.6	36.7			34.7					± 3.0
	g/l	366	367			347					± 30
	mmol/l	22.72	22.81			21.58					± 1.86
IDR RDW	%	15.0	13.5			15.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	74	79			78					± 22
VPM MPV	µm ³ ; fl	9.1	8.5			8.3					± 2.0
LYMPHO	#	1.26	1.27			1.27					± 0.40
	%	63.0	63.6			63.5					± 8.0
MONO	#	0.20	0.25			0.19					± 0.19
	%	10.0	12.5			9.5					± 6.0
GRANULO	#	0.54	0.48			0.54					± 0.40
	%	27.0	24.0			27.0					± 7.0

LOT MX 434 N
Rev 1

CONTROL N

 (Exp.)

2022-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.8			7.6					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.61			4.55					± 0.26
HB HGB	g/dl	13.5	13.6			13.3					± 0.7
	g/l	135	136			133					± 7
	mmol/l	8.38	8.45			8.28					± 0.43
HT HCT	%	37.3	37.3			37.3					± 2.7
	l/l	0.373	0.373			0.373					± 0.027
VGM MCV	µm ³ ; fl	81	81			82					± 4
TGMH MCH	pg	29.3	29.5			29.3					± 2.0
	fmol	1.82	1.83			1.82					± 0.12
CCMH MCHC	g/dl	36.2	36.4			35.7					± 3.0
	g/l	362	364			357					± 30
	mmol/l	22.50	22.62			22.18					± 1.86
IDR RDW	%	15.0	13.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	252	260			258					± 40
VPM MPV	µm ³ ; fl	8.5	7.9			7.8					± 2.0
LYMPHO	#	2.62	2.81			2.55					± 0.40
	%	34.0	36.0			33.5					± 5.0
MONO	#	0.50	0.56			0.49					± 0.30
	%	6.5	7.1			6.5					± 4.0
GRANULO	#	4.58	4.43			4.56					± 0.60
	%	59.5	56.8			60.0					± 6.0

Ref: TEMP-0830 Rev.41 FRONT / RECTO 130060626



LOT

MX 434 H

CONTROL

H

(Exp.)

2022-06-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.2	20.2			19.9					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.73	5.66			5.69					± 0.32
HB HGB	g/dl	18.5	18.3			18.1					± 0.9
	g/l	185	183			181					± 9
	mmol/l	11.49	11.36			11.25					± 0.56
HT HCT	%	50.4	50.9			51.4					± 3.4
	l/l	0.504	0.509			0.514					± 0.034
VGM MCV	µm ³ ; fl	88	90			90					± 4
TGMH MCH	pg	32.3	32.3			31.8					± 2.0
	fmol	2.00	2.01			1.98					± 0.12
CCMH MCHC	g/dl	36.7	35.9			35.4					± 3.0
	g/l	367	359			354					± 30
	mmol/l	22.78	22.31			21.97					± 1.86
IDR RDW	%	13.5	13.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	506	515			515					± 65
VPM MPV	µm ³ ; fl	8.5	7.9			7.8					± 2.0
LYMPHO	#	3.33	3.43			3.28					± 1.00
	%	16.5	17.0			16.5					± 5.0
MONO	#	0.91	1.20			0.90					± 0.80
	%	4.5	5.9			4.5					± 4.0
GRANULO	#	15.96	15.54			15.72					± 1.40
	%	79.0	76.9			79.0					± 7.0