

LOT MX 415 L
 Rev 1

CONTROL L

(Exp.)

2019-04-05
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.37	2.32			2.35					± 0.15
HB HGB	g/dl	5.9	5.7			5.9					± 0.4
	g/l	59	57			59					± 4
	mmol/l	3.66	3.54			3.66					± 0.25
HT HCT	%	16.6	16.2			16.5					± 2.0
	l/l	0.166	0.162			0.165					± 0.020
VGM MCV	µm ³ ; fl	70	70			70					± 4
TGMH MCH	pg	24.9	24.6			25.1					± 2.0
	fmol	1.55	1.53			1.56					± 0.12
CCMH MCHC	g/dl	35.6	35.1			35.9					± 3.0
	g/l	356	351			359					± 30
	mmol/l	22.08	21.80			22.27					± 1.86
IDR RDW	%	14.0	14.6			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	73	64			75					± 20
VPM MPV	µm ³ ; fl	8.8	8.2			8.6					± 2.0
LYMPHO	#	1.31	1.29			1.31					± 0.40
	%	65.5	64.7			65.5					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.5	8.1			8.0					± 6.0
GRANULO	#	0.52	0.54			0.53					± 0.40
	%	26.0	27.2			26.5					± 7.0

LOT

MX 415 N
Rev 1

CONTROL N

(Exp.)

2019-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.7			7.9					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.75	4.65			4.73					± 0.18
HB HGB	g/dl	13.5	13.2			13.5					± 0.5
	g/l	135	132			135					± 5
	mmol/l	8.38	8.20			8.38					± 0.31
HT HCT	%	37.5	37.2			37.4					± 2.5
	l/l	0.375	0.372			0.374					± 0.025
VGM MCV	µm ³ ; fl	79	80			79					± 4
TGMH MCH	pg	28.4	28.4			28.5					± 2.0
	fmol	1.76	1.76			1.77					± 0.12
CCMH MCHC	g/dl	36.0	35.5			36.1					± 3.0
	g/l	360	355			361					± 30
	mmol/l	22.34	22.04			22.44					± 1.86
IDR RDW	%	14.5	14.9			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	255	256			265					± 40
VPM MPV	µm ³ ; fl	8.6	8.1			8.2					± 2.0
LYMPHO	#	2.77	2.76			2.80					± 0.40
	%	36.0	35.9			35.5					± 5.0
MONO	#	0.46	0.50			0.51					± 0.30
	%	6.0	6.5			6.5					± 4.0
GRANULO	#	4.47	4.44			4.58					± 0.60
	%	58.0	57.6			58.0					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626

LOT MX 415 H
Rev 1

CONTROL H

 (Exp.)

2019-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.1	20.1			20.6					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.75	5.65			5.72					± 0.20
HB HGB	g/dl	18.4	17.9			18.4					± 0.6
	g/l	184	179			184					± 6
	mmol/l	11.43	11.12			11.43					± 0.37
HT HCT	%	51.2	50.9			50.9					± 3.0
	l/l	0.512	0.509			0.509					± 0.030
VGM MCV	µm ³ ; fl	89	90			89					± 4
TGMH MCH	pg	32.0	31.7			32.2					± 2.0
	fmol	1.99	1.97			2.00					± 0.12
CCMH MCHC	g/dl	36.0	35.2			36.1					± 3.0
	g/l	360	352			361					± 30
	mmol/l	22.33	21.86			22.45					± 1.86
IDR RDW	%	13.5	14.8			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	510	518			515					± 45
VPM MPV	µm ³ ; fl	8.3	7.7			8.0					± 2.0
LYMPHO	#	3.42	3.34			3.40					± 1.00
	%	17.0	16.6			16.5					± 5.0
MONO	#	0.90	0.96			0.93					± 0.80
	%	4.5	4.8			4.5					± 4.0
GRANULO	#	15.78	15.80			16.27					± 1.40
	%	78.5	78.6			79.0					± 7.0