

LOT		MX 402 L	CONTROL L				(Exp.)				2017-02-05 (YYYY-MM-DD)
Rev 1		Milysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	1.9	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.42	2.42			2.43					± 0.15
HB HGB	g/dl	6.0	6.2			6.0					± 0.4
	g/l	60	62			60					± 4
	mmol/l	3.73	3.83			3.73					± 0.25
HT HCT	%	16.9	16.7			16.8					± 2.0
	l/l	0.169	0.167			0.168					± 0.020
VGM MCV	µm ³ ; fl	70	69			69					± 4
TGMH MCH	pg	24.8	25.5			24.7					± 2.0
	fmol	1.54	1.58			1.53					± 0.12
CCMH MCHC	g/dl	35.4	37.0			35.8					± 3.0
	g/l	354	370			358					± 30
	mmol/l	22.00	22.96			22.22					± 1.86
IDR RDW	%	14.0	14.0			14.4					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	66			73					± 20
VPM MPV	µm ³ ; fl	8.3	8.3			8.0					± 2.0
LYMPHO	#	1.24	1.27			1.30					± 0.40
	%	65.5	63.6			65.0					± 8.0
MONO	#	0.17	0.24			0.20					± 0.20
	%	8.8	12.0			10.0					± 6.0
GRANULO	#	0.49	0.49			0.50					± 0.40
	%	25.7	24.5			25.0					± 7.0



LOT

MX 402 N
Rev 1

CONTROL N

(Exp.)

2017-02-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.4	7.5			7.6					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.78	4.75			4.78					± 0.18
HB HGB	g/dl	13.9	13.7			13.7					± 0.5
	g/l	139	137			137					± 5
	mmol/l	8.63	8.51			8.51					± 0.31
HT HCT	%	38.2	38.0			38.2					± 2.5
	l/l	0.382	0.380			0.382					± 0.025
VGM MCV	µm ³ ; fl	80	80			80					± 4
TGMH MCH	pg	29.1	28.8			28.7					± 2.0
	fmol	1.81	1.79			1.78					± 0.12
CCMH MCHC	g/dl	36.3	36.1			35.8					± 3.0
	g/l	363	361			358					± 30
	mmol/l	22.57	22.39			22.25					± 1.86
IDR RDW	%	13.8	13.5			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	256	252			257					± 40
VPM MPV	µm ³ ; fl	8.5	8.5			8.3					± 2.0
LYMPHO	#	2.55	2.58			2.58					± 0.40
	%	34.5	34.5			34.0					± 5.0
MONO	#	0.48	0.50			0.46					± 0.30
	%	6.5	6.7			6.0					± 4.0
GRANULO	#	4.37	4.41			4.56					± 0.60
	%	59.0	58.8			60.0					± 6.0

1300027266-A
Ref: TEMP-0830 Rev.36



LOT

MX 402 H

CONTROL

H

(Exp.)

2017-02-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.4	20.1			20.3					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.90	5.86			5.91					± 0.20
HB HGB	g/dl	18.7	18.4			18.6					± 0.6
	g/l	187	184			186					± 6
	mmol/l	11.61	11.43			11.55					± 0.37
HT HCT	%	51.9	51.6			51.4					± 3.0
	l/l	0.519	0.516			0.514					± 0.030
VGM MCV	µm ³ ; fl	88	88			87					± 4
TGMH MCH	pg	31.7	31.4			31.5					± 2.0
	fmol	1.97	1.95			1.95					± 0.12
CCMH MCHC	g/dl	36.0	35.7			36.2					± 3.0
	g/l	360	357			362					± 30
	mmol/l	22.37	22.16			22.46					± 1.86
IDR RDW	%	13.5	13.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	515	510			515					± 45
VPM MPV	µm ³ ; fl	7.9	8.0			7.8					± 2.0
LYMPHO	#	3.33	3.09			3.35					± 1.00
	%	16.3	15.4			16.5					± 5.0
MONO	#	0.94	1.19			0.91					± 0.80
	%	4.6	5.9			4.5					± 4.0
GRANULO	#	16.14	15.76			16.04					± 1.40
	%	79.1	78.4			79.0					± 7.0