

**LOT** MX 425 L  
Rev 1

**CONTROL** L

 (Exp.)

**2020-12-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1	2.1			2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.24	2.23			2.18					± 0.15
HB HGB	g/dl	5.8	5.8			5.6					± 0.4
	g/l	58	58			56					± 4
	mmol/l	3.60	3.59			3.48					± 0.25
HT HCT	%	16.1	15.8			15.9					± 2.0
	l/l	0.161	0.158			0.159					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	72	71			73					± 4
TGMH MCH	pg	25.9	25.9			25.7					± 2.0
	fmol	1.61	1.61			1.60					± 0.12
CCMH MCHC	g/dl	36.0	36.5			35.2					± 3.0
	g/l	360	365			352					± 30
	mmol/l	22.33	22.68			21.85					± 1.86
IDR RDW	%	14.5	13.5			14.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	74	77			75					± 20
VPM MPV	µm <sup>3</sup> ; fl	8.5	8.0			8.6					± 2.0
LYMPHO	#	1.38	1.40			1.30					± 0.40
	%	65.5	66.7			65.0					± 8.0
MONO	#	0.20	0.23			0.20					± 0.20
	%	8.5	11.0			8.5					± 6.0
GRANULO	#	0.55	0.47			0.53					± 0.40
	%	26.0	22.5			26.5					± 7.0

**LOT**

**MX 425 N**  
Rev 1

**CONTROL N**

(Exp.)

**2020-12-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>8.0</b>	<b>7.9</b>			<b>7.8</b>					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>4.68</b>	<b>4.66</b>			<b>4.58</b>					± 0.18
HB HGB	g/dl	<b>13.5</b>	<b>13.3</b>			<b>13.3</b>					± 0.5
	g/l	<b>135</b>	<b>133</b>			<b>133</b>					± 5
	mmol/l	<b>8.38</b>	<b>8.26</b>			<b>8.26</b>					± 0.31
HT HCT	%	<b>37.4</b>	<b>36.8</b>			<b>36.6</b>					± 2.5
	l/l	<b>0.374</b>	<b>0.368</b>			<b>0.366</b>					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	<b>80</b>	<b>79</b>			<b>80</b>					± 4
TGMH MCH	pg	<b>28.8</b>	<b>28.5</b>			<b>29.0</b>					± 2.0
	fmol	<b>1.79</b>	<b>1.77</b>			<b>1.80</b>					± 0.12
CCMH MCHC	g/dl	<b>36.1</b>	<b>36.1</b>			<b>36.3</b>					± 3.0
	g/l	<b>361</b>	<b>361</b>			<b>363</b>					± 30
	mmol/l	<b>22.39</b>	<b>22.44</b>			<b>22.54</b>					± 1.86
IDR RDW	%	<b>14.0</b>	<b>13.5</b>			<b>14.0</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>245</b>	<b>246</b>			<b>240</b>					± 40
VPM MPV	µm <sup>3</sup> ; fl	<b>8.3</b>	<b>8.0</b>			<b>8.4</b>					± 2.0
LYMPHO	#	<b>2.87</b>	<b>2.89</b>			<b>2.69</b>					± 0.40
	%	<b>35.9</b>	<b>36.6</b>			<b>34.5</b>					± 5.0
MONO	#	<b>0.50</b>	<b>0.53</b>			<b>0.47</b>					± 0.30
	%	<b>6.2</b>	<b>6.7</b>			<b>6.0</b>					± 4.0
GRANULO	#	<b>4.63</b>	<b>4.49</b>			<b>4.64</b>					± 0.60
	%	<b>57.9</b>	<b>56.8</b>			<b>59.5</b>					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626



**LOT**

**MX 425 H**  
Rev 1

**CONTROL H**

(Exp.)

**2020-12-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>21.0</b>	<b>20.2</b>			<b>20.6</b>					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.59</b>	<b>5.56</b>			<b>5.50</b>					± 0.20
HB HGB	g/dl	<b>18.2</b>	<b>17.6</b>			<b>18.0</b>					± 0.6
	g/l	<b>182</b>	<b>176</b>			<b>180</b>					± 6
	mmol/l	<b>11.30</b>	<b>10.93</b>			<b>11.18</b>					± 0.37
HT HCT	%	<b>49.8</b>	<b>49.5</b>			<b>49.5</b>					± 3.0
	l/l	<b>0.498</b>	<b>0.495</b>			<b>0.495</b>					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>89</b>			<b>90</b>					± 4
TGMH MCH	pg	<b>32.6</b>	<b>31.7</b>			<b>32.7</b>					± 2.0
	fmol	<b>2.02</b>	<b>1.97</b>			<b>2.03</b>					± 0.12
CCMH MCHC	g/dl	<b>36.6</b>	<b>35.6</b>			<b>36.4</b>					± 3.0
	g/l	<b>366</b>	<b>356</b>			<b>364</b>					± 30
	mmol/l	<b>22.72</b>	<b>22.09</b>			<b>22.58</b>					± 1.86
IDR RDW	%	<b>13.5</b>	<b>12.6</b>			<b>13.5</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>497</b>	<b>497</b>			<b>498</b>					± 45
VPM MPV	µm <sup>3</sup> ; fl	<b>8.1</b>	<b>7.7</b>			<b>8.1</b>					± 2.0
LYMPHO	#	<b>3.65</b>	<b>3.43</b>			<b>3.50</b>					± 1.00
	%	<b>17.4</b>	<b>17.0</b>			<b>17.0</b>					± 5.0
MONO	#	<b>0.90</b>	<b>1.20</b>			<b>0.82</b>					± 0.80
	%	<b>4.3</b>	<b>5.9</b>			<b>4.0</b>					± 4.0
GRANULO	#	<b>16.44</b>	<b>15.54</b>			<b>16.27</b>					± 1.40
	%	<b>78.3</b>	<b>76.9</b>			<b>79.0</b>					± 7.0