

**HORIBA ABX SAS**  
HORIBA ABX SAS  
Parc Euromédecine  
Rue du Caducée - BP 7290  
34184 MONTPELLIER Cedex 4  
FRANCE  
Tél.: 33 (0) 4 67 14 15 16  
Fax.: 33 (0) 4 67 14 15 17

LOT		MX 435 L		CONTROL		L		(Exp.)			2022-08-05			
		Rev 1		Minilysebio										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS		MICROS		MICROS		ARGOS - HELIOS		PENTRA		TOLERANCES TOLERANCE
		ES60	CRP 200					60 BP - 60 18P						
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.9	1.9					1.9						± 0.4
GR RBC	10 <sup>12</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.20	2.23					2.22						± 0.17
HB HGB	g/dl	5.7	5.8					5.5						± 0.5
	g/l	57	58					55						± 5
	mmol/l	3.54	3.59					3.41						± 0.25
HT HCT	%	15.8	16.1					16.2						± 2.0
	l/l	0.158	0.161					0.162						± 0.020
VGM MCV	µm <sup>3</sup> ; fl	72	72					73						± 4
TGMH MCH	pg	25.9	25.9					24.7						± 2.0
	fmol	1.61	1.61					1.53						± 0.12
CCMH MCHC	g/dl	36.0	36.0					33.9						± 3.0
	g/l	360	360					339						± 30
	mmol/l	22.35	22.36					21.03						± 1.86
IDR RDW	%	15.0	14.0					14.0						± 3.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	78	81					80						± 22
VPM MPV	µm <sup>3</sup> ; fl	9.0	8.7					8.2						± 2.0
LYMPHO	#	1.24	1.26					1.24						± 0.40
	%	65.0	66.2					65.5						± 8.0
MONO	#	0.18	0.20					0.16						± 0.20
	%	9.5	10.5					8.5						± 6.0
GRANULO	#	0.48	0.45					0.49						± 0.40
	%	25.5	23.5					26.0						± 7.0

LOT		MX 435 N		CONTROL		N		(Exp.)			2022-08-05			
		Rev 1		Minilysebio										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS		MICROS		MICROS		ARGOS - HELIOS		PENTRA		TOLERANCES TOLERANCE
		ES60	CRP 200					60 BP - 60 18P						
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.5	7.4					7.4						± 0.8
GR RBC	10 <sup>12</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.70	4.67					4.63						± 0.26
HB HGB	g/dl	13.6	13.7					13.4						± 0.7
	g/l	136	137					134						± 7
	mmol/l	8.45	8.51					8.34						± 0.31
HT HCT	%	37.6	37.8					38.0						± 2.7
	l/l	0.376	0.378					0.380						± 0.027
VGM MCV	µm <sup>3</sup> ; fl	80	81					82						± 4
TGMH MCH	pg	28.9	29.3					29.0						± 2.0
	fmol	1.80	1.82					1.80						± 0.12
CCMH MCHC	g/dl	36.2	36.2					35.4						± 3.0
	g/l	362	362					354						± 30
	mmol/l	22.46	22.49					21.96						± 1.86
IDR RDW	%	14.0	13.5					14.0						± 3.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	250	257					259						± 40
VPM MPV	µm <sup>3</sup> ; fl	8.5	8.1					7.8						± 2.0
LYMPHO	#	2.66	2.86					2.63						± 0.40
	%	35.5	38.7					35.5						± 5.0
MONO	#	0.49	0.53					0.52						± 0.30
	%	6.5	7.1					7.0						± 4.0
GRANULO	#	4.35	4.02					4.26						± 0.60
	%	58.0	54.4					57.5						± 6.0

LOT		MX 435 H		CONTROL		H		(Exp.)			2022-08-05			
		Rev 1		Minilysebio										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS		MICROS		MICROS		ARGOS - HELIOS		PENTRA		TOLERANCES TOLERANCE
		ES60	CRP 200					60 BP - 60 18P						
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	19.5	19.4					19.5						± 1.7
GR RBC	10 <sup>12</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.68	5.68					5.67						± 0.32
HB HGB	g/dl	18.4	18.2					18.0						± 0.9
	g/l	184	182					180						± 9
	mmol/l	11.43	11.30					11.19						± 0.37
HT HCT	%	50.6	50.6					51.0						± 3.4
	l/l	0.506	0.506					0.510						± 0.034
VGM MCV	µm <sup>3</sup> ; fl	89	89					90						± 4
TGMH MCH	pg	32.4	32.0					31.8						± 2.0
	fmol	2.01	1.99					1.97						± 0.12
CCMH MCHC	g/dl	36.4	36.0					35.3						± 3.0
	g/l	364	360					353						± 30
	mmol/l	22.60	22.36					21.93						± 1.86
IDR RDW	%	13.5	13.0					13.5						± 3.0
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	500	505					510						± 65
VPM MPV	µm <sup>3</sup> ; fl	8.1	7.7					7.5						± 2.0
LYMPHO	#	3.32	3.39					3.12						± 1.00
	%	17.0	17.5					16.0						± 5.0
MONO	#	0.88	1.15					0.88						± 0.80
	%	4.5	5.9					4.5						± 4.0
GRANULO	#	15.31	14.83					15.50						± 1.40
	%	78.5	76.4					79.5						± 7.0