

LOT MX 406 L
Rev 1

CONTROL L

 (Exp.)

2017-10-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.44	2.45			2.40					± 0.15
HB HGB	g/dl	6.1	6.2			6.1					± 0.4
	g/l	61	62			61					± 4
	mmol/l	3.79	3.83			3.79					± 0.25
HT HCT	%	17.1	17.2			16.8					± 2.0
	l/l	0.171	0.172			0.168					± 0.020
VGM MCV	µm ³ ; fl	70	70			70					± 4
TGMH MCH	pg	25.0	25.2			25.4					± 2.0
	fmol	1.55	1.56			1.58					± 0.12
CCMH MCHC	g/dl	35.7	36.0			36.3					± 3.0
	g/l	357	360			363					± 30
	mmol/l	22.18	22.36			22.55					± 1.86
IDR RDW	%	14.0	14.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	75			77					± 20
VPM MPV	µm ³ ; fl	8.8	8.8			8.5					± 2.0
LYMPHO	#	1.32	1.33			1.33					± 0.40
	%	66.0	66.7			66.5					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.0	8.5			7.5					± 6.0
GRANULO	#	0.52	0.50			0.52					± 0.40
	%	26.0	25.0			26.0					± 7.0

LOT MX 406 N
 Rev 1

CONTROL N

(Exp.)

2017-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.4	7.4			7.5					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.82	4.77			4.78					± 0.18
HB HGB	g/dl	13.5	13.4			13.5					± 0.5
	g/l	135	134			135					± 5
	mmol/l	8.38	8.32			8.38					± 0.31
HT HCT	%	39.0	38.2			37.8					± 2.5
	l/l	0.390	0.382			0.378					± 0.025
VGM MCV	µm ³ ; fl	81	80			79					± 4
TGMH MCH	pg	28.0	28.1			28.2					± 2.0
	fmol	1.74	1.74			1.75					± 0.12
CCMH MCHC	g/dl	34.6	35.1			35.8					± 3.0
	g/l	346	351			358					± 30
	mmol/l	21.47	21.81			22.20					± 1.86
IDR RDW	%	14.5	13.8			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	254	250			251					± 40
VPM MPV	µm ³ ; fl	8.4	8.3			8.0					± 2.0
LYMPHO	#	2.55	2.59			2.63					± 0.40
	%	34.5	35.0			35.0					± 5.0
MONO	#	0.52	0.59			0.45					± 0.30
	%	7.0	8.0			6.0					± 4.0
GRANULO	#	4.33	4.21			4.43					± 0.60
	%	58.5	56.8			59.0					± 6.0

1300027266-A
 Ref: TEMP-0830 Rev.38



LOT MX 406 H
 Rev 1

CONTROL H

(Exp.)

2017-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.0	20.1			20.3					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.80	5.76			5.77					± 0.20
HB HGB	g/dl	18.4	18.0			18.4					± 0.6
	g/l	184	180			184					± 6
	mmol/l	11.43	11.18			11.43					± 0.37
HT HCT	%	51.6	50.7			50.8					± 3.0
	l/l	0.516	0.507			0.508					± 0.030
VGM MCV	µm ³ ; fl	89	88			88					± 4
TGMH MCH	pg	31.7	31.3			31.9					± 2.0
	fmol	1.97	1.94			1.98					± 0.12
CCMH MCHC	g/dl	35.6	35.5			36.2					± 3.0
	g/l	356	355			362					± 30
	mmol/l	22.14	22.05			22.50					± 1.86
IDR RDW	%	13.5	13.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	501	500			500					± 45
VPM MPV	µm ³ ; fl	7.5	7.4			7.2					± 2.0
LYMPHO	#	3.30	3.30			3.45					± 1.00
	%	16.5	16.4			17.0					± 5.0
MONO	#	0.90	1.41			0.91					± 0.80
	%	4.5	7.0			4.5					± 4.0
GRANULO	#	15.80	15.36			15.94					± 1.40
	%	79.0	76.4			78.5					± 7.0