

**LOT** MX 426 L  
 Rev 1

**CONTROL** L

(Exp.)

**2021-02-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 BP - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.0			2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.43	2.38			2.40					± 0.15
HB HGB	g/dl	5.9	5.8			5.7					± 0.4
	g/l	59	58			57					± 4
	mmol/l	3.66	3.59			3.54					± 0.25
HT HCT	%	16.8	16.7			17.0					± 2.0
	l/l	0.168	0.167			0.170					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	69	70			71					± 4
TGMH MCH	pg	24.3	24.3			23.8					± 2.0
	fmol	1.51	1.51			1.47					± 0.12
CCMH MCHC	g/dl	35.2	34.7			33.5					± 3.0
	g/l	352	347			335					± 30
	mmol/l	21.85	21.55			20.77					± 1.86
IDR RDW	%	14.5	13.8			15.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	74			72					± 20
VPM MPV	µm <sup>3</sup> ; fl	8.4	8.2			8.4					± 2.0
LYMPHO	#	1.32	1.34			1.31					± 0.40
	%	66.0	66.8			65.5					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	7.0	9.1			7.0					± 6.0
GRANULO	#	0.54	0.49			0.55					± 0.40
	%	27.0	24.3			27.5					± 7.0

**LOT** MX 426 N  
Rev 1


**CONTROL** N

 (Exp.)

**2021-02-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.7	7.3			7.5					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.66	4.56			4.64					± 0.18
HB HGB	g/dl	13.7	13.5			13.5					± 0.5
	g/l	137	135			135					± 5
	mmol/l	8.51	8.38			8.38					± 0.31
HT HCT	%	38.7	37.8			38.5					± 2.5
	l/l	0.387	0.378			0.385					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	83	83			83					± 4
TGMH MCH	pg	29.4	29.6			29.1					± 2.0
	fmol	1.83	1.84			1.81					± 0.12
CCMH MCHC	g/dl	35.4	35.7			35.1					± 3.0
	g/l	354	357			351					± 30
	mmol/l	22.00	22.15			21.77					± 1.86
IDR RDW	%	14.5	13.5			14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	250	248			253					± 40
VPM MPV	µm <sup>3</sup> ; fl	7.6	7.4			7.7					± 2.0
LYMPHO	#	2.81	2.81			2.70					± 0.40
	%	36.5	38.5			36.0					± 5.0
MONO	#	0.46	0.51			0.45					± 0.30
	%	6.0	7.0			6.0					± 4.0
GRANULO	#	4.43	3.99			4.35					± 0.60
	%	57.5	54.7			58.0					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626

<b>LOT</b>		<b>MX 426 H</b>		<b>CONTROL H</b>						 (Exp.)		<b>2021-02-05</b> (YYYY-MM-DD)				
		Rev 1														
		<b>Minilysebio</b>														
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS		MICROS		MICROS		MICROS		ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200					60 8P - 60 18P								
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.2</b>	<b>19.5</b>					<b>19.9</b>								± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.76</b>	<b>5.67</b>					<b>5.75</b>								± 0.20
HB HGB	g/dl	<b>18.7</b>	<b>18.0</b>					<b>18.5</b>								± 0.6
	g/l	<b>187</b>	<b>180</b>					<b>185</b>								± 6
	mmol/l	<b>11.61</b>	<b>11.18</b>					<b>11.49</b>								± 0.37
HT HCT	%	<b>52.4</b>	<b>51.6</b>					<b>51.8</b>								± 3.0
	l/l	<b>0.524</b>	<b>0.516</b>					<b>0.518</b>								± 0.030
VGM MCV	µm <sup>3</sup> ; fl	<b>91</b>	<b>91</b>					<b>90</b>								± 4
TGMH MCH	pg	<b>32.5</b>	<b>31.7</b>					<b>32.2</b>								± 2.0
	fmol	<b>2.02</b>	<b>1.97</b>					<b>2.00</b>								± 0.12
CCMH MCHC	g/dl	<b>35.7</b>	<b>34.9</b>					<b>35.7</b>								± 3.0
	g/l	<b>357</b>	<b>349</b>					<b>357</b>								± 30
	mmol/l	<b>22.15</b>	<b>21.66</b>					<b>22.20</b>								± 1.86
IDR RDW	%	<b>13.5</b>	<b>12.8</b>					<b>13.5</b>								± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>509</b>	<b>505</b>					<b>522</b>								± 45
VPM MPV	µm <sup>3</sup> ; fl	<b>7.5</b>	<b>7.2</b>					<b>7.6</b>								± 2.0
LYMPHO	#	<b>3.43</b>	<b>3.41</b>					<b>3.28</b>								± 1.00
	%	<b>17.0</b>	<b>17.5</b>					<b>16.5</b>								± 5.0
MONO	#	<b>0.81</b>	<b>1.07</b>					<b>0.80</b>								± 0.80
	%	<b>4.0</b>	<b>5.5</b>					<b>4.0</b>								± 4.0
GRANULO	#	<b>15.96</b>	<b>14.98</b>					<b>15.82</b>								± 1.40
	%	<b>79.0</b>	<b>76.8</b>					<b>79.5</b>								± 7.0