

**LOT** MX 417 L  
 Rev 1

**CONTROL** L

(Exp.)

**2019-08-05**  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	1.9	1.8			2.0					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.31	2.28			2.27					± 0.15
HB HGB	g/dl	5.8	5.7			5.7					± 0.4
	g/l	58	57			57					± 4
	mmol/l	3.60	3.53			3.54					± 0.25
HT HCT	%	16.2	16.0			15.9					± 2.0
	l/l	0.162	0.160			0.159					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	70	70			70					± 4
TGMH MCH	pg	25.1	24.9			25.1					± 2.0
	fmol	1.56	1.55			1.56					± 0.12
CCMH MCHC	g/dl	35.9	35.6			35.9					± 3.0
	g/l	359	356			359					± 30
	mmol/l	22.27	22.12			22.28					± 1.86
IDR RDW	%	14.0	13.5			14.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	75	80			72					± 20
VPM MPV	µm <sup>3</sup> ; fl	9.5	9.0			9.1					± 2.0
LYMPHO	#	1.28	1.24			1.34					± 0.40
	%	67.5	68.7			67.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	7.5	9.0			7.5					± 6.0
GRANULO	#	0.48	0.41			0.51					± 0.40
	%	25.0	22.5			25.5					± 7.0

**LOT** MX 417 N  
Rev 1

**CONTROL** N

 (Exp.)

**2019-08-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.7	7.3			7.8					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.62	4.57			4.58					± 0.18
HB HGB	g/dl	13.5	13.3			13.4					± 0.5
	g/l	135	133			134					± 5
	mmol/l	8.38	8.26			8.32					± 0.31
HT HCT	%	37.4	37.0			37.1					± 2.5
	l/l	0.374	0.370			0.371					± 0.025
VGM MCV	µm <sup>3</sup> ; fl	81	81			81					± 4
TGMH MCH	pg	29.2	29.1			29.3					± 2.0
	fmol	1.81	1.81			1.82					± 0.12
CCMH MCHC	g/dl	36.1	35.9			36.1					± 3.0
	g/l	361	359			361					± 30
	mmol/l	22.40	22.31			22.43					± 1.86
IDR RDW	%	13.5	13.5			14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	243	253			254					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.3	7.8			8.0					± 2.0
LYMPHO	#	2.73	2.67			2.69					± 0.40
	%	35.5	36.6			34.5					± 5.0
MONO	#	0.54	0.52			0.51					± 0.30
	%	7.0	7.1			6.5					± 4.0
GRANULO	#	4.43	4.11			4.60					± 0.60
	%	57.5	56.4			59.0					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626

**LOT** MX 417 H  
Rev 1

**CONTROL** H

 (Exp.)

**2019-08-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	19.7	19.4			20.2					± 1.6
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.64	5.56			5.58					± 0.20
HB HGB	g/dl	18.3	17.6			18.3					± 0.6
	g/l	183	176			183					± 6
	mmol/l	11.36	10.93			11.36					± 0.37
HT HCT	%	50.2	49.5			49.7					± 3.0
	l/l	0.502	0.495			0.497					± 0.030
VGM MCV	µm <sup>3</sup> ; fl	89	89			89					± 4
TGMH MCH	pg	32.4	31.7			32.8					± 2.0
	fmol	2.01	1.97			2.04					± 0.12
CCMH MCHC	g/dl	36.5	35.6			36.8					± 3.0
	g/l	365	356			368					± 30
	mmol/l	22.64	22.09			22.88					± 1.86
IDR RDW	%	13.0	13.0			13.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	500	515			520					± 45
VPM MPV	µm <sup>3</sup> ; fl	8.4	8.0			8.3					± 2.0
LYMPHO	#	3.25	3.29			3.23					± 1.00
	%	16.5	17.0			16.0					± 5.0
MONO	#	0.89	1.15			0.91					± 0.80
	%	4.5	5.9			4.5					± 4.0
GRANULO	#	15.56	14.92			16.06					± 1.40
	%	79.0	76.9			79.5					± 7.0