

LOT MX 427 L
Rev 1

CONTROL L



2021-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.32			2.33					± 0.15
HB HGB	g/dl	5.9	5.8			5.7					± 0.4
	g/l	59	58			57					± 4
	mmol/l	3.66	3.59			3.54					± 0.25
HT HCT	%	16.6	16.0			16.3					± 2.0
	l/l	0.166	0.160			0.163					± 0.020
VGM MCV	µm ³ ; fl	69	69			70					± 4
TGMH MCH	pg	24.6	24.9			24.5					± 2.0
	fmol	1.53	1.55			1.52					± 0.12
CCMH MCHC	g/dl	35.6	36.1			34.9					± 3.0
	g/l	356	361			349					± 30
	mmol/l	22.13	22.43			21.70					± 1.86
IDR RDW	%	14.5	14.0			15.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	83	83			83					± 20
VPM MPV	µm ³ ; fl	8.8	8.1			8.8					± 2.0
LYMPHO	#	1.37	1.35			1.32					± 0.40
	%	65.0	67.7			66.0					± 8.0
MONO	#	0.17	0.19			0.15					± 0.20
	%	8.0	9.5			7.5					± 6.0
GRANULO	#	0.57	0.46			0.53					± 0.40
	%	27.0	23.0			26.5					± 7.0

LOT MX 427 N
Rev 1

CONTROL N

 (Exp.)

2021-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.8	7.5			7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.76	4.67			4.70					± 0.18
HB HGB	g/dl	13.7	13.5			13.7					± 0.5
	g/l	137	135			137					± 5
	mmol/l	8.51	8.38			8.51					± 0.31
HT HCT	%	38.1	37.8			38.1					± 2.5
	l/l	0.381	0.378			0.381					± 0.025
VGM MCV	µm ³ ; fl	80	81			81					± 4
TGMH MCH	pg	28.8	28.9			29.1					± 2.0
	fmol	1.79	1.80			1.81					± 0.12
CCMH MCHC	g/dl	36.0	35.7			36.0					± 3.0
	g/l	360	357			360					± 30
	mmol/l	22.34	22.16			22.35					± 1.86
IDR RDW	%	14.0	13.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	268	274			275					± 40
VPM MPV	µm ³ ; fl	8.3	8.1			8.7					± 2.0
LYMPHO	#	2.77	2.83			2.73					± 0.40
	%	35.5	37.7			35.5					± 5.0
MONO	#	0.51	0.55			0.49					± 0.30
	%	6.5	7.3			6.4					± 4.0
GRANULO	#	4.52	4.13			4.47					± 0.60
	%	58.0	55.1			58.1					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626

LOT MX 427 H
Rev 1

CONTROL H

 (Exp.)

2021-04-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	21.0	20.3			20.8					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.74	5.65			5.73					± 0.20
HB HGB	g/dl	18.4	17.7			18.3					± 0.6
	g/l	184	177			183					± 6
	mmol/l	11.43	10.99			11.36					± 0.37
HT HCT	%	50.5	49.7			50.4					± 3.0
	l/l	0.505	0.497			0.504					± 0.030
VGM MCV	µm ³ ; fl	88	88			88					± 4
TGMH MCH	pg	32.1	31.3			31.9					± 2.0
	fmol	1.99	1.95			1.98					± 0.12
CCMH MCHC	g/dl	36.4	35.6			36.3					± 3.0
	g/l	364	356			363					± 30
	mmol/l	22.62	22.11			22.54					± 1.86
IDR RDW	%	13.0	12.8			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	527	536			545					± 45
VPM MPV	µm ³ ; fl	8.1	7.8			8.5					± 2.0
LYMPHO	#	3.57	3.46			3.45					± 1.00
	%	17.0	17.1			16.6					± 5.0
MONO	#	0.95	1.16			0.87					± 0.80
	%	4.5	5.7			4.2					± 4.0
GRANULO	#	16.49	15.64			16.47					± 1.40
	%	78.5	77.0			79.2					± 7.0