

LOT MX 418 L
 Rev 1

CONTROL L

(Exp.)

2019-10-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	1.9			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.36	2.29			2.29					± 0.15
HB HGB	g/dl	5.9	5.9			5.8					± 0.4
	g/l	59	59			58					± 4
	mmol/l	3.66	3.65			3.60					± 0.25
HT HCT	%	16.5	16.0			16.0					± 2.0
	l/l	0.165	0.160			0.160					± 0.020
VGM MCV	µm ³ ; fl	70	70			70					± 4
TGMH MCH	pg	25.0	25.7			25.3					± 2.0
	fmol	1.55	1.59			1.57					± 0.12
CCMH MCHC	g/dl	35.7	36.7			36.2					± 3.0
	g/l	357	367			362					± 30
	mmol/l	22.18	22.78			22.47					± 1.86
IDR RDW	%	14.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	78			78					± 20
VPM MPV	µm ³ ; fl	8.6	8.1			8.7					± 2.0
LYMPHO	#	1.32	1.29			1.32					± 0.40
	%	66.0	67.7			66.0					± 8.0
MONO	#	0.20	0.18			0.20					± 0.20
	%	8.0	9.5			8.0					± 6.0
GRANULO	#	0.52	0.44			0.52					± 0.40
	%	26.0	23.0			26.0					± 7.0

LOT MX 418 N
Rev 1

CONTROL N

 (Exp.)

2019-10-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 BP - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.8	7.3			7.8					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.65	4.55			4.59					± 0.18
HB HGB	g/dl	13.5	13.2			13.3					± 0.5
	g/l	135	132			133					± 5
	mmol/l	8.38	8.20			8.26					± 0.31
HT HCT	%	37.2	36.4			36.7					± 2.5
	l/l	0.372	0.364			0.367					± 0.025
VGM MCV	µm ³ ; fl	80	80			80					± 4
TGMH MCH	pg	29.0	29.0			29.0					± 2.0
	fmol	1.80	1.80			1.80					± 0.12
CCMH MCHC	g/dl	36.3	36.3			36.2					± 3.0
	g/l	363	363			362					± 30
	mmol/l	22.54	22.52			22.49					± 1.86
IDR RDW	%	14.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	248	253			262					± 40
VPM MPV	µm ³ ; fl	8.3	7.9			8.4					± 2.0
LYMPHO	#	2.89	2.82			2.85					± 0.40
	%	37.0	38.7			36.5					± 5.0
MONO	#	0.51	0.49			0.47					± 0.30
	%	6.5	6.7			6.0					± 4.0
GRANULO	#	4.41	4.01			4.49					± 0.60
	%	56.5	54.9			57.5					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626



LOT

MX 418 H
Rev 1

CONTROL H

(Exp.)

2019-10-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.4	20.0			20.9					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.66	5.57			5.59					± 0.20
HB HGB	g/dl	18.3	17.5			18.1					± 0.6
	g/l	183	175			181					± 6
	mmol/l	11.36	10.87			11.24					± 0.37
HT HCT	%	50.4	49.6			49.8					± 3.0
	l/l	0.504	0.496			0.498					± 0.030
VGM MCV	µm ³ ; fl	89	89			89					± 4
TGMH MCH	pg	32.3	31.4			32.4					± 2.0
	fmol	2.01	1.95			2.01					± 0.12
CCMH MCHC	g/dl	36.3	35.3			36.4					± 3.0
	g/l	363	353			364					± 30
	mmol/l	22.56	21.92			22.59					± 1.86
IDR RDW	%	13.5	13.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	508	515			532					± 45
VPM MPV	µm ³ ; fl	8.1	7.8			8.2					± 2.0
LYMPHO	#	3.67	3.60			3.55					± 1.00
	%	18.0	18.0			17.0					± 5.0
MONO	#	1.02	1.30			1.05					± 0.80
	%	5.0	6.5			5.0					± 4.0
GRANULO	#	15.71	15.09			16.30					± 1.40
	%	77.0	75.5			78.0					± 7.0