



LOT

MX 428 L
Rev 1

CONTROL L



2021-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.0			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.40	2.34			2.36					± 0.15
HB HGB	g/dl	6.0	5.8			5.9					± 0.4
	g/l	60	58			59					± 4
	mmol/l	3.73	3.59			3.66					± 0.25
HT HCT	%	16.6	15.9			16.5					± 2.0
	l/l	0.166	0.159			0.165					± 0.020
VGM MCV	µm ³ ; fl	69	68			70					± 4
TGMH MCH	pg	25.0	24.7			25.0					± 2.0
	fmol	1.55	1.53			1.55					± 0.12
CCMH MCHC	g/dl	36.2	36.3			35.7					± 3.0
	g/l	362	363			357					± 30
	mmol/l	22.50	22.57			22.18					± 1.86
IDR RDW	%	14.5	14.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	81			70					± 20
VPM MPV	µm ³ ; fl	9.0	8.6			9.2					± 2.0
LYMPHO	#	1.43	1.41			1.44					± 0.40
	%	68.0	70.7			68.5					± 6.0
MONO	#	0.20	0.18			0.20					± 0.20
	%	7.0	9.0			7.0					± 5.0
GRANULO	#	0.53	0.41			0.51					± 0.40
	%	25.0	20.6			24.5					± 6.0

LOT MX 428 N
Rev 1

CONTROL N

 (Exp.)

2021-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.5			7.8					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.66	4.60			4.64					± 0.18
HB HGB	g/dl	13.6	13.2			13.4					± 0.5
	g/l	136	132			134					± 5
	mmol/l	8.45	8.20			8.32					± 0.31
HT HCT	%	37.3	36.8			37.6					± 2.5
	l/l	0.373	0.368			0.376					± 0.025
VGM MCV	µm ³ ; fl	80	80			81					± 4
TGMH MCH	pg	29.2	28.7			28.9					± 2.0
	fmol	1.81	1.78			1.79					± 0.12
CCMH MCHC	g/dl	36.5	35.9			35.7					± 3.0
	g/l	365	359			357					± 30
	mmol/l	22.65	22.28			22.14					± 1.86
IDR RDW	%	14.0	13.8			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	244	248			250					± 40
VPM MPV	µm ³ ; fl	8.4	8.1			8.6					± 2.0
LYMPHO	#	2.70	2.78			2.73					± 0.40
	%	35.0	37.1			35.0					± 5.0
MONO	#	0.50	0.53			0.47					± 0.30
	%	6.5	7.1			6.0					± 4.0
GRANULO	#	4.50	4.19			4.60					± 0.60
	%	58.5	55.9			59.0					± 6.0

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626

LOT MX 428 H
Rev 1

CONTROL H

 (Exp.)

2021-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.6	19.8			20.8					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.67	5.55			5.65					± 0.20
HB HGB	g/dl	18.3	17.5			18.2					± 0.6
	g/l	183	175			182					± 6
	mmol/l	11.36	10.87			11.30					± 0.37
HT HCT	%	49.9	48.8			50.3					± 3.0
	l/l	0.499	0.488			0.503					± 0.030
VGM MCV	µm ³ ; fl	88	88			89					± 4
TGMH MCH	pg	32.3	31.5			32.2					± 2.0
	fmol	2.00	1.96			2.00					± 0.12
CCMH MCHC	g/dl	36.7	35.8			36.2					± 3.0
	g/l	367	358			362					± 30
	mmol/l	22.78	22.25			22.48					± 1.86
IDR RDW	%	13.0	13.2			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	475	490			490					± 45
VPM MPV	µm ³ ; fl	8.4	8.1			8.6					± 2.0
LYMPHO	#	3.61	3.36			3.64					± 1.00
	%	17.5	17.0			17.5					± 5.0
MONO	#	0.93	1.18			0.94					± 0.80
	%	4.5	5.9			4.5					± 4.0
GRANULO	#	16.07	15.23			16.22					± 1.40
	%	78.0	76.9			78.0					± 7.0