



LOT

MX 429 L
Rev 1

CONTROL L

(Exp.)

2021-08-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.0			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.25	2.24			2.26					± 0.15
HB HGB	g/dl	5.9	5.9			5.8					± 0.4
	g/l	59	59			58					± 4
	mmol/l	3.66	3.65			3.60					± 0.25
HT HCT	%	16.4	16.4			16.7					± 2.0
	l/l	0.164	0.164			0.167					± 0.020
VGM MCV	µm ³ ; fl	73	73			74					± 4
TGMH MCH	pg	26.2	26.3			25.7					± 2.0
	fmol	1.63	1.63			1.59					± 0.12
CCMH MCHC	g/dl	35.9	36.0			34.7					± 3.0
	g/l	359	360			347					± 30
	mmol/l	22.31	22.33			21.54					± 1.86
IDR RDW	%	15.2	14.0			15.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	70	70			73					± 20
VPM MPV	µm ³ ; fl	8.3	8.1			8.8					± 2.0
LYMPHO	#	1.37	1.39			1.35					± 0.40
	%	68.5	69.7			67.5					± 8.0
MONO	#	0.20	0.16			0.20					± 0.20
	%	6.0	8.0			6.5					± 6.0
GRANULO	#	0.51	0.45			0.52					± 0.40
	%	25.5	22.5			26.0					± 7.0

LOT		MX 429 N	CONTROL N								(Exp.)	2021-08-05 (YYYY-MM-DD)
Rev 1		Minilysebio										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		ES60	CRP 200			60 8P - 60 18P						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.6	7.6			7.8					± 0.8	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.50	4.50			4.50					± 0.18	
HB HGB	g/dl	13.5	13.3			13.4					± 0.5	
	g/l	135	133			134					± 5	
	mmol/l	8.38	8.26			8.32					± 0.31	
HT HCT	%	37.4	37.4			37.8					± 2.5	
	l/l	0.374	0.374			0.378					± 0.025	
VGM MCV	µm ³ ; fl	83	83			84					± 4	
TGMH MCH	pg	30.0	29.6			29.8					± 2.0	
	fmol	1.86	1.84			1.85					± 0.12	
CCMH MCHC	g/dl	36.1	35.6			35.4					± 3.0	
	g/l	361	356			354					± 30	
	mmol/l	22.45	22.11			22.01					± 1.86	
IDR RDW	%	13.5	13.5			14.0					± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	235	243			242					± 40	
VPM MPV	µm ³ ; fl	8.1	7.9			9.0					± 2.0	
LYMPHO	#	2.74	2.86			2.89					± 0.40	
	%	36.0	37.6			37.0					± 5.0	
MONO	#	0.53	0.57			0.55					± 0.30	
	%	7.0	7.6			7.0					± 4.0	
GRANULO	#	4.33	4.17			4.37					± 0.60	
	%	57.0	54.9			56.0					± 6.0	

Ref: TEMP-0830 Rev.39 FRONT / RECTO 130060626

LOT

MX 429 H
Rev 1

CONTROL H

 (Exp.)

2021-08-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.1	20.2			20.5					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.58	5.55			5.59					± 0.20
HB HGB	g/dl	18.4	17.9			18.3					± 0.6
	g/l	184	179			183					± 6
	mmol/l	11.43	11.12			11.36					± 0.37
HT HCT	%	50.2	50.5			51.4					± 3.0
	l/l	0.502	0.505			0.514					± 0.030
VGM MCV	µm ³ ; fl	90	91			92					± 4
TGMH MCH	pg	33.0	32.3			32.7					± 2.0
	fmol	2.05	2.00			2.03					± 0.12
CCMH MCHC	g/dl	36.6	35.4			35.6					± 3.0
	g/l	366	354			356					± 30
	mmol/l	22.75	22.01			22.10					± 1.86
IDR RDW	%	14.0	13.5			14.2					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	465	480			485					± 45
VPM MPV	µm ³ ; fl	8.0	7.8			8.8					± 2.0
LYMPHO	#	3.32	3.53			3.38					± 1.00
	%	16.5	17.5			16.5					± 5.0
MONO	#	0.90	1.20			0.92					± 0.80
	%	4.5	5.9			4.5					± 4.0
GRANULO	#	15.88	15.44			16.20					± 1.40
	%	79.0	76.4			79.0					± 7.0