

LOT MX 412 L
Rev 1

CONTROL L

 (Exp.)

2018-10-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-------------|---------|--------|--------|----------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 | CRP 200 | | | 60 8P - 60 18P | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.1 | 2.1 | | | 2.1 | | | | | ± 0.4 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.27 | 2.28 | | | 2.27 | | | | | ± 0.15 |
| HB HGB | g/dl | 5.8 | 5.9 | | | 5.8 | | | | | ± 0.4 |
| | g/l | 58 | 59 | | | 58 | | | | | ± 4 |
| | mmol/l | 3.60 | 3.65 | | | 3.60 | | | | | ± 0.25 |
| HT HCT | % | 16.1 | 16.2 | | | 16.1 | | | | | ± 2.0 |
| | l/l | 0.161 | 0.162 | | | 0.161 | | | | | ± 0.020 |
| VGM MCV | µm ³ ; fl | 71 | 71 | | | 71 | | | | | ± 4 |
| TGMH MCH | pg | 25.6 | 25.8 | | | 25.6 | | | | | ± 2.0 |
| | fmol | 1.59 | 1.60 | | | 1.59 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.0 | 36.3 | | | 36.0 | | | | | ± 3.0 |
| | g/l | 360 | 363 | | | 360 | | | | | ± 30 |
| | mmol/l | 22.35 | 22.56 | | | 22.35 | | | | | ± 1.86 |
| IDR RDW | % | 14.0 | 13.5 | | | 14.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 77 | 80 | | | 80 | | | | | ± 20 |
| VPM MPV | µm ³ ; fl | 8.8 | 8.3 | | | 8.4 | | | | | ± 2.0 |
| LYMPHO | # | 1.39 | 1.38 | | | 1.38 | | | | | ± 0.40 |
| | % | 66.0 | 65.7 | | | 65.5 | | | | | ± 8.0 |
| MONO | # | 0.20 | 0.19 | | | 0.20 | | | | | ± 0.20 |
| | % | 8.0 | 9.0 | | | 7.5 | | | | | ± 6.0 |
| GRANULO | # | 0.55 | 0.54 | | | 0.57 | | | | | ± 0.40 |
| | % | 26.0 | 25.5 | | | 27.0 | | | | | ± 7.0 |

LOT MX 412 N
Rev 1

CONTROL N

 (Exp.)

2018-10-05
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|-------------|---------|--------|--------|----------------|----------------|--------|--------|--------|-------------------------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | PENTRA | |
| | | ES60 | CRP 200 | | | 60 8P - 60 18P | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 7.3 | 7.4 | | | 7.6 | | | | | ± 0.8 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 4.58 | 4.57 | | | 4.58 | | | | | ± 0.18 |
| HB HGB | g/dl | 13.3 | 13.3 | | | 13.3 | | | | | ± 0.5 |
| | g/l | 133 | 133 | | | 133 | | | | | ± 5 |
| | mmol/l | 8.26 | 8.26 | | | 8.26 | | | | | ± 0.31 |
| HT HCT | % | 36.6 | 36.6 | | | 36.6 | | | | | ± 2.5 |
| | l/l | 0.366 | 0.366 | | | 0.366 | | | | | ± 0.025 |
| VGM MCV | µm ³ ; fl | 80 | 80 | | | 80 | | | | | ± 4 |
| TGMH MCH | pg | 29.0 | 29.1 | | | 29.0 | | | | | ± 2.0 |
| | fmol | 1.80 | 1.81 | | | 1.80 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.3 | 36.4 | | | 36.3 | | | | | ± 3.0 |
| | g/l | 363 | 364 | | | 363 | | | | | ± 30 |
| | mmol/l | 22.54 | 22.59 | | | 22.54 | | | | | ± 1.86 |
| IDR RDW | % | 14.0 | 13.5 | | | 14.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 262 | 265 | | | 265 | | | | | ± 40 |
| VPM MPV | µm ³ ; fl | 8.4 | 7.8 | | | 7.9 | | | | | ± 2.0 |
| LYMPHO | # | 2.48 | 2.59 | | | 2.62 | | | | | ± 0.40 |
| | % | 34.0 | 35.0 | | | 34.5 | | | | | ± 5.0 |
| MONO | # | 0.47 | 0.53 | | | 0.46 | | | | | ± 0.30 |
| | % | 6.5 | 7.1 | | | 6.0 | | | | | ± 4.0 |
| GRANULO | # | 4.34 | 4.28 | | | 4.52 | | | | | ± 0.60 |
| | % | 59.5 | 57.8 | | | 59.5 | | | | | ± 6.0 |

1300027266-A
Ref: TEMP-0830 Rev.36

| LOT | | MX 412 H | CONTROL H | | | | (Exp.) | | 2018-10-05 (YYYY-MM-DD) | | |
|--------------------------|--|--------------------|------------------|--------|--------|----------------|----------------|--------|-----------------------------------|--------|-------------------------|
| Rev 1 | | Minilysebio | | | | | | | | | |
| PARAMETRES PARAMETERS | UNITES UNITS | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | PENTRA | TOLERANCES TOLERANCE |
| | | ES60 | CRP 200 | | | 60 8P - 60 18P | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 20.1 | 20.4 | | | 20.5 | | | | | ± 1.6 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 5.65 | 5.60 | | | 5.60 | | | | | ± 0.20 |
| HB HGB | g/dl | 18.4 | 17.8 | | | 18.2 | | | | | ± 0.6 |
| | g/l | 184 | 178 | | | 182 | | | | | ± 6 |
| | mmol/l | 11.43 | 11.05 | | | 11.30 | | | | | ± 0.37 |
| HT HCT | % | 50.3 | 49.8 | | | 49.8 | | | | | ± 3.0 |
| | l/l | 0.503 | 0.498 | | | 0.498 | | | | | ± 0.030 |
| VGM MCV | µm ³ ; fl | 89 | 89 | | | 89 | | | | | ± 4 |
| TGMH MCH | pg | 32.6 | 31.8 | | | 32.5 | | | | | ± 2.0 |
| | fmol | 2.02 | 1.97 | | | 2.02 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.6 | 35.7 | | | 36.5 | | | | | ± 3.0 |
| | g/l | 366 | 357 | | | 365 | | | | | ± 30 |
| | mmol/l | 22.72 | 22.18 | | | 22.68 | | | | | ± 1.86 |
| IDR RDW | % | 13.5 | 13.0 | | | 13.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 520 | 525 | | | 524 | | | | | ± 45 |
| VPM MPV | µm ³ ; fl | 8.5 | 7.9 | | | 8.0 | | | | | ± 2.0 |
| LYMPHO | # | 3.12 | 3.14 | | | 3.28 | | | | | ± 1.00 |
| | % | 15.5 | 15.4 | | | 16.0 | | | | | ± 5.0 |
| MONO | # | 0.90 | 1.32 | | | 1.03 | | | | | ± 0.80 |
| | % | 4.5 | 6.5 | | | 5.0 | | | | | ± 4.0 |
| GRANULO | # | 16.08 | 15.89 | | | 16.20 | | | | | ± 1.40 |
| | % | 80.0 | 77.9 | | | 79.0 | | | | | ± 7.0 |