

LOT MX 431 L
 Rev 1

CONTROL L

(Exp.)

2021-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.2	2.1			2.2					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.26	2.22			2.25					± 0.15
HB HGB	g/dl	5.8	5.8			5.6					± 0.4
	g/l	58	58			56					± 4
	mmol/l	3.60	3.59			3.48					± 0.25
HT HCT	%	16.0	15.8			16.2					± 2.0
	l/l	0.160	0.158			0.162					± 0.020
VGM MCV	µm ³ ; fl	71	71			72					± 4
TGMH MCH	pg	25.7	26.0			24.9					± 2.0
	fmol	1.59	1.62			1.55					± 0.12
CCMH MCHC	g/dl	36.1	36.7			34.6					± 3.0
	g/l	361	367			346					± 30
	mmol/l	22.45	22.78			21.47					± 1.86
IDR RDW	%	15.0	13.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	78			77					± 20
VPM MPV	µm ³ ; fl	9.2	9.0			8.5					± 2.0
LYMPHO	#	1.45	1.40			1.45					± 0.40
	%	66.0	66.7			66.0					± 8.0
MONO	#	0.15	0.21			0.18					± 0.15
	%	7.0	10.0			8.0					± 6.0
GRANULO	#	0.59	0.49			0.57					± 0.40
	%	27.0	23.5			26.0					± 7.0

LOT

MX 431 N

CONTROL

N

(Exp.)

2021-12-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.8	7.7			7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.58	4.50			4.58					± 0.18
HB HGB	g/dl	13.7	13.4			13.3					± 0.5
	g/l	137	134			133					± 5
	mmol/l	8.51	8.32			8.28					± 0.31
HT HCT	%	37.6	36.9			38.0					± 2.5
	l/l	0.376	0.369			0.380					± 0.025
VGM MCV	µm ³ ; fl	82	82			83					± 4
TGMH MCH	pg	29.9	29.8			29.1					± 2.0
	fmol	1.86	1.85			1.81					± 0.12
CCMH MCHC	g/dl	36.5	36.3			35.1					± 3.0
	g/l	365	363			351					± 30
	mmol/l	22.65	22.55			21.77					± 1.86
IDR RDW	%	14.0	13.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	253	260			266					± 40
VPM MPV	µm ³ ; fl	8.7	8.4			8.0					± 2.0
LYMPHO	#	2.73	2.82			2.73					± 0.40
	%	35.0	36.6			34.5					± 5.0
MONO	#	0.59	0.62			0.55					± 0.30
	%	7.5	8.0			7.0					± 4.0
GRANULO	#	4.49	4.26			4.62					± 0.60
	%	57.5	55.4			58.5					± 6.0

Ref: TEMP-0830 Rev.41 FRONT / RECTO 130060626

LOT MX 431 H
Rev 1

CONTROL H

 (Exp.)

2021-12-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.1	19.7			19.7					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.87	5.84			5.86					± 0.20
HB HGB	g/dl	18.5	18.1			18.5					± 0.6
	g/l	185	181			185					± 6
	mmol/l	11.49	11.24			11.49					± 0.37
HT HCT	%	51.7	51.4			51.7					± 3.0
	l/l	0.517	0.514			0.517					± 0.030
VGM MCV	µm ³ ; fl	88	88			88					± 4
TGMH MCH	pg	31.5	31.0			31.6					± 2.0
	fmol	1.96	1.92			1.96					± 0.12
CCMH MCHC	g/dl	35.8	35.2			35.9					± 3.0
	g/l	358	352			359					± 30
	mmol/l	22.24	21.87			22.28					± 1.86
IDR RDW	%	13.5	13.5			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	480	495			497					± 45
VPM MPV	µm ³ ; fl	8.0	7.8			7.8					± 2.0
LYMPHO	#	3.52	3.24			3.25					± 1.00
	%	16.5	16.4			16.0					± 5.0
MONO	#	0.80	1.06			0.81					± 0.80
	%	4.5	5.4			4.0					± 4.0
GRANULO	#	15.78	15.35			16.24					± 1.40
	%	79.0	77.9			80.0					± 7.0