



LOT MX 016 L
 Rev 1

CONTROL L

(Exp.)

2016-04-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.15
HB HGB	g/dl		6.0			5.6					± 0.4
	g/l		60			56					± 4
	mmol/l		3.71			3.47					± 0.25
HT HCT	%		16.5			16.5					± 2.0
	l/l		0.165			0.165					± 0.020
VGM MCV	µm ³ ; fl		68			68					± 4
TGMH MCH	pg		24.6			23.1					± 2.0
	fmol		1.53			1.43					± 0.12
CCMH MCHC	g/dl		36.2			33.9					± 3.0
	g/l		362			339					± 30
	mmol/l		22.47			21.06					± 1.86
IDR RDW	%		14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		75			75					± 20
VPM MPV	µm ³ ; fl		8.8			8.5					± 2.0
LYMPHO	#		1.36			1.31					± 0.40
	%		68.2			65.5					± 8.0
MONO	#		0.19			0.20					± 0.20
	%		9.5			10.0					± 6.0
GRANULO	#		0.45			0.49					± 0.40
	%		22.5			24.5					± 7.0

LOT MX 016 N
 Rev 1

CONTROL N

(Exp.)

2016-04-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.18
HB HGB	g/dl		7.2			7.2					± 0.5
	g/l		13.4			13.1					± 5
	mmol/l		8.32			8.15					± 0.31
HT HCT	%		37.0			37.1					± 2.5
	l/l		0.370			0.371					± 0.025
VGM MCV	µm ³ ; fl		80			80					± 4
TGMH MCH	pg		29.0			28.3					± 2.0
	fmol		1.80			1.76					± 0.12
CCMH MCHC	g/dl		36.3			35.4					± 3.0
	g/l		363			354					± 30
	mmol/l		22.51			21.97					± 1.86
IDR RDW	%		13.5			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		253			256					± 40
VPM MPV	µm ³ ; fl		8.2			7.8					± 2.0
LYMPHO	#		2.48			2.42					± 0.50
	%		34.5			33.7					± 6.0
MONO	#		0.54			0.53					± 0.40
	%		7.6			7.4					± 5.0
GRANULO	#		4.16			4.24					± 0.60
	%		57.8			59.0					± 7.0

1219930092-A
 Ref: TEMP-0830 Rev.37



LOT

MX 016 H

CONTROL

H

(Exp.)

2016-04-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l		CRP 200			60 8P - 60 18P					± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l										± 0.20
HB HGB	g/dl		18.2			17.9					± 0.6
	g/l		182			179					± 6
	mmol/l		11.30			11.13					± 0.37
HT HCT	%		50.0			50.4					± 3.0
	l/l		0.500			0.504					± 0.030
VGM MCV	µm ³ ; fl		89			89					± 4
TGMH MCH	pg		32.4			31.7					± 2.0
	fmol		2.01			1.97					± 0.12
CCMH MCHC	g/dl		36.4			35.6					± 3.0
	g/l		364			356					± 30
	mmol/l		22.60			22.09					± 1.86
IDR RDW	%		13.0			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l		503			502					± 45
VPM MPV	µm ³ ; fl		8.0			7.7					± 2.0
LYMPHO	#		3.12			3.07					± 1.00
	%		15.9			16.0					± 5.0
MONO	#		1.27			1.06					± 0.80
	%		6.5			5.5					± 4.0
GRANULO	#		15.17			15.02					± 1.40
	%		77.4			78.2					± 7.0