



LOT MX 035 L
 Rev 1

CONTROL L

(Exp.)

2015-06-05
 (YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-------------|--------|--------|--------|--------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | | 2.0 | | | 2.1 | | | | | ± 0.4 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | | 2.35 | | | 2.34 | | | | | ± 0.15 |
| HB HGB | g/dl | | 6.0 | | | 5.7 | | | | | ± 0.4 |
| | g/l | | 60 | | | 57 | | | | | ± 4 |
| | mmol/l | | 3.71 | | | 3.52 | | | | | ± 0.25 |
| HT HCT | % | | 16.2 | | | 16.1 | | | | | ± 2.0 |
| | l/l | | 0.162 | | | 0.161 | | | | | ± 0.020 |
| VGM MCV | µm ³ ; fl | | 69 | | | 69 | | | | | ± 4 |
| TGMH MCH | pg | | 25.4 | | | 24.2 | | | | | ± 2.0 |
| | fmol | | 1.58 | | | 1.51 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | | 36.9 | | | 35.1 | | | | | ± 3.0 |
| | g/l | | 369 | | | 351 | | | | | ± 30 |
| | mmol/l | | 22.89 | | | 21.82 | | | | | ± 1.86 |
| IDR RDW | % | | 13.5 | | | 13.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | | 74 | | | 75 | | | | | ± 20 |
| VPM MPV | µm ³ ; fl | | 9.0 | | | 8.5 | | | | | ± 2.0 |
| LYMPHO | # | | 1.25 | | | 1.28 | | | | | ± 0.40 |
| | % | | 62.6 | | | 61.0 | | | | | ± 8.0 |
| MONO | # | | 0.22 | | | 0.23 | | | | | ± 0.20 |
| | % | | 11.0 | | | 11.0 | | | | | ± 6.0 |
| GRANULO | # | | 0.53 | | | 0.59 | | | | | ± 0.40 |
| | % | | 26.5 | | | 28.0 | | | | | ± 7.0 |

LOT MX 035 N
 Rev 1

CONTROL N

(Exp.)

2015-06-05
 (YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-------------|--------|--------|--------|--------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | | 7.9 | | | 7.7 | | | | | ± 0.8 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | | 4.55 | | | 4.55 | | | | | ± 0.18 |
| HB HGB | g/dl | | 13.5 | | | 13.2 | | | | | ± 0.5 |
| | g/l | | 135 | | | 132 | | | | | ± 5 |
| | mmol/l | | 8.38 | | | 8.22 | | | | | ± 0.31 |
| HT HCT | % | | 36.9 | | | 36.9 | | | | | ± 2.5 |
| | l/l | | 0.369 | | | 0.369 | | | | | ± 0.025 |
| VGM MCV | µm ³ ; fl | | 81 | | | 81 | | | | | ± 4 |
| TGMH MCH | pg | | 29.7 | | | 29.1 | | | | | ± 2.0 |
| | fmol | | 1.84 | | | 1.81 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | | 36.6 | | | 35.9 | | | | | ± 3.0 |
| | g/l | | 366 | | | 359 | | | | | ± 30 |
| | mmol/l | | 22.75 | | | 22.29 | | | | | ± 1.86 |
| IDR RDW | % | | 13.0 | | | 13.3 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | | 259 | | | 262 | | | | | ± 40 |
| VPM MPV | µm ³ ; fl | | 8.3 | | | 8.1 | | | | | ± 2.0 |
| LYMPHO | # | | 2.60 | | | 2.42 | | | | | ± 0.40 |
| | % | | 32.9 | | | 31.6 | | | | | ± 5.0 |
| MONO | # | | 0.53 | | | 0.50 | | | | | ± 0.30 |
| | % | | 6.7 | | | 6.5 | | | | | ± 4.0 |
| GRANULO | # | | 4.76 | | | 4.75 | | | | | ± 0.60 |
| | % | | 60.3 | | | 62.0 | | | | | ± 6.0 |

9930002-A
 Ref: TEMP-0830 Rev.36



LOT

MX 035 H
 Rev 1

CONTROL H

(Exp.)

2015-06-05
 (YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-------------|---------|--------|--------|----------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | | CRP 200 | | | 60 8P - 60 18P | | | | | ± 1.6 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | | | | | | | | | | ± 0.20 |
| HB HGB | g/dl | | 18.2 | | | 18.1 | | | | | ± 0.6 |
| | g/l | | 182 | | | 181 | | | | | ± 6 |
| | mmol/l | | 11.30 | | | 11.25 | | | | | ± 0.37 |
| HT HCT | % | | 51.5 | | | 51.7 | | | | | ± 3.0 |
| | l/l | | 0.515 | | | 0.517 | | | | | ± 0.030 |
| VGM MCV | µm ³ ; fl | | 87 | | | 87 | | | | | ± 4 |
| TGMH MCH | pg | | 30.7 | | | 30.5 | | | | | ± 2.0 |
| | fmol | | 1.91 | | | 1.89 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | | 35.3 | | | 35.1 | | | | | ± 3.0 |
| | g/l | | 353 | | | 351 | | | | | ± 30 |
| | mmol/l | | 21.94 | | | 21.77 | | | | | ± 1.86 |
| IDR RDW | % | | 12.5 | | | 13.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | | 500 | | | 510 | | | | | ± 45 |
| VPM MPV | µm ³ ; fl | | 8.2 | | | 8.0 | | | | | ± 2.0 |
| LYMPHO | # | | 3.53 | | | 3.33 | | | | | ± 1.00 |
| | % | | 17.0 | | | 16.5 | | | | | ± 5.0 |
| MONO | # | | 1.35 | | | 1.11 | | | | | ± 0.80 |
| | % | | 6.5 | | | 5.5 | | | | | ± 4.0 |
| GRANULO | # | | 15.90 | | | 15.68 | | | | | ± 1.40 |
| | % | | 76.4 | | | 77.7 | | | | | ± 7.0 |